

Aptima assayvätskesats (Panther™-systemet)

För *in vitro*-diagnostiskt bruk.

Rx only

Avsedd användning

Aptima assayvätskesats är avsedd för användning på Panther-systemet.

Tillhandahållen materiel

Aptima assayvätskesats (kat.nr 303014)

(Förvaras vid 15 °C till 30 °C efter leverans)

Symbol	Komponent	Antal
W	Aptima tvättlösning <i>10 mM HEPES-buffrad lösning som innehåller mindre än 2 % konserveringsmedel.</i>	1 x 2,9 L
O	Aptima oljereagens <i>Silikonolja.</i>	1 x 260 mL
DF	Aptima buffert för inaktiveringsvätska <i>800 mM bikarbonatbuffrad lösning.</i>	1 x 1,4 L

Varningar och försiktighetsåtgärder

- A. Iaktta sedvanliga säkerhetsrutiner för laboratorier.
- B. Undvik att kontaminera Aptima assayvätskor med mikrober och ribonukleas.

Krav på förvaring och hantering

- A. Förvara öppnade Aptima assayvätskor vid 15 °C till 30 °C.
- B. När de öppnats är Aptima assayvätskor stabila i 60 dagar.
- C. Använd inte satsen efter utgångsdatum.
- D. Vissa reagenser i denna sats är märkta med faro- och säkerhetssymboler.

Obs! För information om farokommunikation, se Safety Data Sheet Library (bibliotek med säkerhetsdatablad) på www.hologicds.com.

USA faroinformation



Buffer For Deactivation Fluid

SODIUM HYDROXIDE 1 – 5%




WARNING

H315 – Causes skin irritation

H319 – Causes serious eye irritation

P264 – Wash face, hands and any exposed skin thoroughly after handling

P280 – Wear protective gloves/protective clothing/eye protection/face protection

	<p>P305 + P351 + P338 – IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing</p> <p>P337 + P313 – If eye irritation persists: Get medical advice/attention</p> <p>P302 + P352 – IF ON SKIN: Wash with plenty of soap and water</p> <p>P332 + P313 – If skin irritation occurs: Get medical advice/attention</p> <p>P362 – Take off contaminated clothing and wash before reuse</p>
	<p>Aptima Oil POLYDIMETHYLSILOXANE 95 – 100% WARNING</p> <p>H315 – Causes skin irritation H319 – Causes serious eye irritation</p> <p>P264 – Wash face, hands and any exposed skin thoroughly after handling</p> <p>P280 – Wear protective gloves/protective clothing/eye protection/face protection</p> <p>P305 + P351 + P338 – IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing</p> <p>P337 + P313 – If eye irritation persists: Get medical advice/attention</p> <p>P302 + P352 – IF ON SKIN: Wash with plenty of soap and water</p> <p>P332 + P313 – If skin irritation occurs: Get medical advice/attention</p> <p>P362 – Take off contaminated clothing and wash before reuse</p>
EU faroinformation	
	<p>Buffer For Deactivation Fluid NATRIUMHYDROXID 1 – 5 % <i>Sodium Hypochlorite <1 %</i> WARNING</p> <p>H315 - Irriterar huden H319 - Orsakar allvarlig ögonirritation</p>
	<p>Aptima Oil POLYDIMETHYLSILOXANE 100 % WARNING</p> <p>H315 - Irriterar huden H319 - Orsakar allvarlig ögonirritation</p>

Procedur

Byt ut flaskorna i universalvätskelådan när systemet uppmanar om detta. Se *Panther-systemets användarmanual*.



Hologic, Inc.
10210 Genetic Center Drive
San Diego, CA 92121 USA

Kundservice: +1 800 442 9892
customersupport@hologic.com

Teknisk support: +1 888 484 4747
molecularsupport@hologic.com

Besök www.hologic.com för ytterligare kontaktinformation.

Hologic, Aptima, och Panther eller registrerade varumärken som tillhör Hologic, Inc. och/eller dess dotterbolag i USA och/eller andra länder. Alla andra varumärken som uppträder i denna bipacksedel är varumärken som tillhör sina respektive ägare.

Denna produkt kan omfattas av ett eller flera USA-patent som identifieras på www.hologic.com/patents.

©2001–2018 Hologic, Inc. Med ensamrätt.

502181SV Rev. 003
2018-01

Hologic N.V.
Da Vincilaan 5
1930 Zaventem
Belgium