

# Aptima® Multitest Swab Specimen Collection Kit for SARS-CoV-2 Specimen Collection

US Only: For Emergency Use Authorization (EUA) Only

**IVD**

## Intended Use

The Aptima® Multitest Swab Specimen Collection Kit is intended to be used for clinician collection of throat and nasal swab specimens for testing with Hologic assays to detect the presence of RNA for SARS-CoV-2.

## Materials Provided

### 50 Multitest Swab Specimen Collection Kits (Cat. No. PRD-03546)

Each kit contains:

Component	Quantity	Description
Swab	1	Individually wrapped, sterile swab
Transport tube	1	Tube containing Specimen Transport Medium (STM), 2.9 mL
<i>Patient collection instructions for vaginal swab, penile meatal swab, rectal swab, and throat swab specimens</i>	1 package	Pack of patient collection instructions for vaginal swab, penile meatal swab, rectal swab, and throat swab specimens

**Note:** The patient collection instructions provided in the Aptima Multitest Swab Specimen Collection Kit do not include instructions for SARS-CoV-2 specimen collection. Please refer to this package insert (AW-18109) for throat and nasal swab specimen collection and handling for SARS-CoV-2 instructions.

## Warnings and Precautions

- A. Use the provided swab only. Failure to use the provided swab may invalidate the test results.
- B. Do not apply the specimen transport medium directly to skin or mucous membranes or take internally.
- C. Specimens may be infectious. Use Universal Precautions when handling specimens. Only personnel adequately trained in handling infectious materials should be permitted to handle specimens.
- D. Take care to avoid cross-contamination during the specimen handling steps. Specimens can contain extremely high levels of pathogens. Ensure that specimen containers do not contact one another, and discard used materials without passing over the containers. If gloves come in contact with specimen, change gloves to avoid cross-contamination.
- E. If the contents of the transport tube are spilled at any time during the collection procedure, use a new Aptima Multitest Swab Specimen Collection Kit. Failure to use a new kit may invalidate the test results.
- F. Maintain proper storage conditions during specimen shipping to ensure the integrity of the specimen. Specimen stability under shipping conditions other than those recommended has not been evaluated.
- G. Do not use the kit after its expiration date to collect specimens.

## Kit Storage Requirements

Store collection kit at room temperature (15°C to 30°C).

## Multitest Swab Specimen Performance

Assay performance characteristics using the Multitest swab specimens are provided in the appropriate package insert for use with Aptima assays and other Hologic products. The Hologic assay package inserts may be referenced online at [www.hologic.com](http://www.hologic.com).

When collecting multiple specimens from the same patient, the tube label provides a specimen source field for unique identification of each specimen.

## Throat Swab Specimens Collection and Handling

Instructions for throat swab specimen collection:

1. Partially peel open the swab package. Remove the swab. Do not touch the soft tip or lay the swab down. If the soft tip is touched, the swab is laid down, or the swab is dropped, use a new Aptima Multitest Swab Specimen Collection Kit.
2. Hold the swab, placing your thumb and forefinger in the middle of the swab shaft covering the score line. Do not hold the swab shaft below the score line.
3. Carefully insert the swab into the throat ensuring contact with bilateral tonsils (if present) and the posterior pharyngeal wall, then withdraw the swab without touching the inside of the cheeks or tongue.
4. While holding the swab in the same hand, unscrew the cap from the tube. Do not spill the contents of the tube. If the contents of the tube are spilled, use a new Aptima Multitest Swab Specimen Collection Kit.
5. Immediately place the swab into the transport tube so that the score line is at the top of the tube.
6. Carefully break the swab shaft at the score line against the side of the tube.
7. Immediately discard the top portion of the swab shaft.
8. Tightly screw the cap onto the tube.

## Nasal Swab Specimens Collection and Handling

Instructions for nasal swab specimen collection:

1. Partially peel open the swab package. Remove the swab. Do not touch the soft tip or lay the swab down. If the soft tip is touched, the swab is laid down, or the swab is dropped, use a new Aptima Multitest Swab Specimen Collection Kit.
2. Hold the swab, placing your thumb and forefinger in the middle of the swab shaft covering the score line. Do not hold the swab shaft below the score line.
3. Carefully insert the swab into the first nostril until resistance is met at the level of the turbinates (less than one inch into the nostril). Rotate the swab a few times against the nasal wall and remove from nostril.
4. Using the same swab, carefully insert the swab into the second nostril until resistance is met at the level of the turbinates (less than one inch into the nostril). Rotate the swab a few times against the nasal wall and remove from nostril.

5. While holding the swab in the same hand, unscrew the cap from the tube. Do not spill the contents of the tube. If the contents of the tube are spilled, use a new Aptima Multitest Swab Specimen Collection Kit.
6. Immediately place the swab into the transport tube so that the score line is at the top of the tube.
7. Carefully break the swab shaft at the score line against the side of the tube.
8. Immediately discard the top portion of the swab shaft.
9. Tightly screw the cap onto the tube.

## Specimen Transport and Storage

Multitest swab specimens are intended to be transported to the laboratory in the provided swab specimen transport medium and tube for use with Hologic SARS-CoV-2 assays. Refer to the appropriate Hologic assay package insert for proper specimen storage conditions.

**Note:** Specimens must be shipped in accordance with applicable national and international regulations.

## Limitations

- A. Use this collection kit only with Aptima assays and other Hologic products. Performance has not been established with other products.
- B. Use appropriate personal protective equipment (PPE) when collecting and handling specimens from individuals suspected of being infected with SARS-CoV-2 as outlined in CDC Interim Laboratory Biosafety Guidelines for Handling and Processing Specimens Associated with 2019 Novel Coronavirus (2019-nCoV).



Hologic, Inc.  
10210 Genetic Center Drive  
San Diego, CA 92121 USA



**Hologic BVBA**  
Da Vincilaan 5  
1930 Zaventem  
Belgium

Customer Support: +1 800 442 9892  
customersupport@hologic.com

Technical Support: +1 888 484 4747  
molecularsupport@hologic.com

For more contact information visit [www.hologic.com](http://www.hologic.com).

Hologic, Aptima, and associated logos, are trademarks and/or registered trademarks of Hologic, Inc. and/or its subsidiaries in the United States and/or other countries.

All other trademarks that may appear in this package insert are the property of their respective owners.

This product may be covered by one or more U.S. patents identified at [www.hologic.com/patents](http://www.hologic.com/patents).

©2016-2020 Hologic, Inc. All rights reserved.

AW-18109-001 Rev. 003  
2020-05