

BKV Primers**Intended Use**

Analyte Specific Reagent (ASR). Analytical and performance characteristics are not established. This ASR is intended to be used as a component of a laboratory developed test (LDT) used exclusively by laboratories. Laboratories may develop their own LDTs using one or more ASRs and/or general laboratory reagents for *in vitro* diagnostic purposes.

Product Description

The BK virus (BKV) Primers ASR contains forward and reverse primers designed to amplify a 111 bp region of the VP2 gene of BKV.

Materials Provided

BKV Primers (Cat. No. PRD-04991)

Component	Quantity	Concentration
Tubes	1 x 0.53 mL	37.5 µmol/L of each primer

Warnings and Precautions

- A. Analytical and performance characteristics are not established.
- B. Use routine laboratory precautions. Wear disposable, powderless gloves, protective eye wear, and laboratory coats when handling specimens and kit reagents. Wash hands thoroughly after handling reagents.
- C. Dispose of all materials that have come in contact with reagents according to local, state, and federal regulations.

Quality Control

This analyte specific reagent was manufactured and released in accordance with the Hologic quality control and quality assurance procedures.

Storage and Handling

Upon receipt, store material at -15 to -85 °C. Multiple freeze/thaw cycles should be minimized by appropriate aliquoting and should not exceed 3 freeze/thaw cycles.



Hologic, Inc.
10210 Genetic Center Drive
San Diego, CA 92121 USA

Customer Support: +1 800 442 9892
customersupport@hologic.com

Technical Support: +1 888 484 4747
molecularsupport@hologic.com

For more contact information visit www.hologic.com.

Hologic and associated logos are trademarks or registered trademarks of Hologic, Inc. and/or its subsidiaries in the United States and/or other countries.

All other trademarks that may appear in this package insert are the property of their respective owners.

This product may be covered by one or more U.S. patents identified at www.hologic.com/patents.

© 2018 Hologic, Inc. All rights reserved.

AW-18448 Rev. 001
2018-12