

FLUOROSCAN[®] InSight[®]-FD



InSight[®] FD Mini C-arm

Customer Release Notes

MAN-09001 Revision 001

HOLOGIC[®]

InSight® FD

Mini C-arm

Customer Release Notes

For Software Version 6.1.4

Part Number MAN-09001

Revision 001

February 2022

Caution: Federal (U.S.A.) law restricts this device to sale by or on the order of a physician (or properly licensed practitioner).

Customer Support

Toll Free in USA: +1.800.321.4659

Email: SkeletalHealth.Support@hologic.com

In Europe, South America, or Asia, contact your local dealer or distributor.

© 2022 Hologic, Inc. Printed in the USA. This manual was originally written in English.

Hologic, FluorSCAN, InSight, MegaView and associated logos are trademarks and/or registered trademarks of Hologic, Inc., and/or its subsidiaries in the United States and/or other countries. All other trademarks, registered trademarks, and product names are the property of their respective owners.

This product may be protected by one or more U.S. or foreign patents as identified at www.Hologic.com/patent-information.



Hologic Inc.
36 Apple Ridge Road
Danbury, CT 06810 USA
1.800.447.1856



Hologic BV
Da Vincilaan 5
1930 Zaventem
Belgium
Tel: +32 2 711 46 80
Fax: +32 2 725 20 87



Brazilian Contact:
Imex Medical Group do Brasil
Rua das Embaúbas, 601
Fazenda Santo Antônio
São José /SC 88104-561 Brasil
+55 48 3251-8800
www.imexmedicalgroup.com.br

HOLOGIC®

Table of Contents

1	InSight® Software Release Notes	1
1.1	InSight 6.1.4	1
1.2	InSight 6.1.3	1
1.3	InSight 6.1.2	1
1.4	InSight 6.1.1	1
1.5	InSight 6.1	2
1.6	InSight 6.0.2	2
1.7	InSight 6.0.1	2
1.8	InSight 6.0	3
1.9	InSight 5.2.1	3
1.10	InSight 5.2	3
1.11	InSight 5.1.1	4
1.12	InSight 5.1	4
1.13	InSight 5.0.6	4
1.14	InSight 5.0.5	4
1.15	InSight 5.0.4	4
1.16	InSight 5.0.3	4
1.17	InSight 5.0.2	5
1.18	InSight 5.0.1	5

InSight® Software Release Notes

1.1 InSight 6.1.4

Added

- DICOM Radiation Dose Structured Report
- Support for FIPS-140-2 compliant Wi-Fi networking
- Support for new image board

Resolved Issues

- Added kV and mA values to standard DICOM attributes
- Deprecated <F1> Windows Help

1.2 InSight 6.1.3

Added

- Display the DAP values in mGy·cm² units
- Audible notification when x-rays are on or stopped

Note

- Hologic only to support use of systems with Instant Off enabled

1.3 InSight 6.1.2

Added

- Support for Windows 10
- Language translation improvements

1.4 InSight 6.1.1

Resolved Issues

- “S” Saved icon not appearing on current acquisitions image
- Unrecoverable error when acquiring images in High Resolution mode
- It was possible to enable High Resolution mode and Auto IQ or Low Dose modes simultaneously by changing modes for the tube head control panel
- It was possible to get invalid mA value in Low Dose mode
- Using Ctrl key to select multiple images on Review screen could lose image selection
- Display UDI in About dialog box
- Issues with displayed dates when date format is changed to German
- German language translation improvements

1.5 InSight 6.1

Features

- Low Dose Rate Imaging mode
- Ability to manually load patient information ahead of time
- Indication when images are saved
- Indication when images are received by PACS
- Auto-Save option
- MegaView™ in review mode
- Improved patient search capability
- Ability to rotate image continuously from tube head
- High Resolution Imaging Mode (unbinned images)
- Easy Zoom
- Improved annotations
- Ability to view only saved images
- Collimator positioning error
- Updated 10-minute dose alarm
- Enhanced GUI buttons

1.6 InSight 6.0.2

Features

- User-adjustable settings for maximum exposure time for x-rays in a session

Resolved Issues

- The patient list may not load properly in non-English language
- The default mode may not display correctly in non-English language
- F3 key or Flip Menu option did not update the status of the button
- Image could have been acquired in select or enter patient screen when session was just restored
- Lost description when interrupted session was restored and allowed selecting another patient
- Updated select language strings

1.7 InSight 6.0.1

Features

- kV; mA; dt; Air Kerma and Air Kerma rate displayed when MegaView Image in use

1.8 InSight 6.0

Features

- Current image highlighted in the acquisition view film strip
- Preview filmstrip add for widescreen monitors
- Windows close control in upper left hand corner removed.
- Images per session increase to 500
- Option for full screen single image added for widescreen monitor
- Number of images stored increased to 10,000
- Cine time increased to 40 minutes
- An option to send DAP report automatically when a study is sent to PACS
- Configuration added to disable warp correction on an II
- SMPTE pattern added to evaluate the monitor image quality
- Check added that system time has not been lost indicating a battery problem
- Warning added if brightness is set to zero
- Cumulative Air Kerma added to DAP report
- User interface updated for widescreen monitor

Resolved Issues

- X-rays disabled if acquisition pedal is held down after the snapshot is complete
- Unprocessed storage commits can cause XRC error messages
- Adding a physician with the language option prevents the added physician from acquiring images
- White speckles displayed when acquiring of metal in reverse video
- DAP report prints light on UP-898 printer
- Live image thumbnail displayed in upper left-hand corner

1.9 InSight 5.2.1

Resolved Issues

- Merge libraries corrupt on install

1.10 InSight 5.2

Features

- User interface adjusted for widescreen monitor
- Imaging firmware updated for motherboard revision
- Wi-Fi drivers updated for Windows 7

1.11 InSight 5.1.1

Resolved Issues

- White halo appears at edge of image circle on InSight 2 systems

1.12 InSight 5.1

Features

- DICOM tag (0008,0033) populated
- Resolved Issues
- DAP data not saved if system unplugged without ending session
- X-rays disabled after 15 minutes of continuous Cine acquisition
- User can acquire image in Select patient screen
- Study description lost if unit is powered off during acquisition
- Study description not displayed on image Acquisition screen
- DAP report does not round time to nearest second
- Unrecoverable error on InSight 2 systems after 10 to 15 Cine loops are acquired

1.13 InSight 5.0.6

Features

- Include Performed Series Sequence, Patient ID, and exposure dose in MPPS N-SET message

1.14 InSight 5.0.5

Features

- Improve flattening algorithm
- Correct black image appearance in Perform
- Remove message box on startup

1.15 InSight 5.0.4

Features

- Reset imaging board at startup

1.16 InSight 5.0.3

Features

- DICOM images have all post processing applied such as image flipping and sharpening before being sent

1.17 InSight 5.0.2

Features

- Reinitialize detector at startup

1.18 InSight 5.0.1

Features

- Dynamically update detector flattening at startup