

GI Expanded Bacterial Assay (Panther Fusion® System)

Instructions for Use

For *in vitro* diagnostic use only

Rx Only

General Information	2
Intended Use	2
Summary and Explanation of the Test	2
Principles of the Procedure	3
Warnings and Precautions	4
Reagent Storage and Handling Requirements	7
Specimen Collection and Storage	8
Specimen Transport	9
Panther Fusion System	10
Reagents and Materials Provided for Panther Fusion GI Expanded Bacterial Assay	10
Materials Required and Available Separately	11
Panther Fusion System Test Procedure	12
Procedural Notes	13
Quality Control	14
Negative and Positive Controls	14
Internal Control	14
Interpretation of Results	15
Limitations	16
Analytical Performance	17
Analytical Sensitivity	17
Inclusivity/Reactivity - Wet Testing	17
Inclusivity/Reactivity - In Silico Analysis	20
Analytical Specificity: Cross Reactivity and Microbial Interference - Wet Testing	20
Coinfection/Competitive Interference	22
Interference	23
Carryover Contamination	26
Reproducibility	27
Clinical Performance	29
Bibliography	32
Contact Information	33

General Information

Intended Use

The Panther Fusion® GI Expanded Bacterial Assay is a multiplex real-time PCR *in vitro* diagnostic test for the rapid and qualitative detection and differentiation of *Yersinia enterocolitica*, *Vibrio* (*V. parahaemolyticus*, *V. vulnificus*, *V. cholerae*), *Escherichia coli* O157, and *Plesiomonas shigelloides*. Nucleic acids are isolated and purified from Cary-Blair preserved stool specimens collected from individuals exhibiting signs and symptoms of gastroenteritis.

This assay is intended to aid in the differential diagnosis of *Yersinia enterocolitica*, *Vibrio* (*V. parahaemolyticus*, *V. vulnificus*, *V. cholerae*), *Escherichia coli* O157, and *Plesiomonas shigelloides* infections. The results of this assay should be used in conjunction with clinical presentation, laboratory findings, and epidemiological information and should not be used as the sole basis for diagnosis, treatment, or other patient management decisions. Positive results do not rule out coinfection with other organisms that are not detected by this test and may not be the sole or definitive cause of patient illness. Negative results in the setting of clinical illness compatible with gastroenteritis may be due to infection by pathogens that are not detected by this test, or non-infectious causes such as ulcerative colitis, irritable bowel syndrome, or Crohn's disease. This assay is designed for use on the Panther Fusion® System.

Summary and Explanation of the Test

Acute diarrhea is a leading cause of outpatient visits, hospitalization, and lost quality of life in both domestic settings and among those traveling abroad. The global impact of foodborne disease is substantial with an estimated 600 million people becoming ill, resulting in 420,000 deaths annually.¹ The Centers for Disease Control and Prevention (CDC) has estimated 48 million cases of foodborne illness annually in the US leading to 128,000 hospitalizations and 3,000 deaths.² Acute diarrhea is associated with estimated healthcare costs upwards of \$150 million.³

Infectious gastroenteritis can be caused by a variety of bacterial, viral, and parasitic organisms. Symptoms alone cannot be used to distinguish the cause of the infection, making rapid and accurate diagnostic tools essential for guiding treatment and patient management.

CDC estimates *Y. enterocolitica* causes 116,716 illnesses, 637 hospitalizations, and 34 deaths in the United States every year.⁴ Children are infected more often than adults, and the infection is more common in the winter.⁵

Vibriosis causes an estimated 80,000 illnesses and 100 deaths in the United States every year. Most infections occur from May through October when water temperatures are warmer.⁶ About 52,000 of these illnesses are estimated to be the result of eating contaminated food.⁶ The most commonly reported species, *V. parahaemolyticus*, is estimated to cause 45,000 illnesses each year in the United States.⁵

An estimated 265,000 Shigatoxin *Escherichia coli* (STEC) infections occur each year in the United States, with STEC O157 causing about 36% of these infections.⁴ Public health experts rely on estimates rather than actual numbers of infections because not all STEC infections are diagnosed.⁷

Outbreaks of diarrheal disease have been associated with contaminated water and oysters containing *P. shigelloides*, and reduction in the severity and duration of symptoms following appropriate antimicrobial therapy has been observed.⁸

Principles of the Procedure

The Panther Fusion System fully automates specimen processing, including sample lysis, nucleic acid capture, amplification, and detection for the Panther Fusion GI Expanded Bacterial Assay. Nucleic acid capture and elution takes place in a single tube on the Panther Fusion System. The eluate is transferred to the Panther Fusion System reaction tube containing the assay reagents. Multiplex real-time PCR is then performed for the eluted nucleic acid on the Panther Fusion System.

Sample processing: Prior to processing and testing on the Panther Fusion System, specimens are transferred to an Aptima® Multitest tube containing specimen transport media (STM) that lyses the cells, releases target nucleic acid, and protects them from degradation during storage.

Nucleic acid capture and elution: An internal control (IC-B) is added automatically to each specimen via the working Panther Fusion Capture Reagent-B (wFCR-B) to monitor for interference during specimen processing, amplification, and detection caused by reagent failure or inhibitory substances. Specimens are first incubated in an alkaline reagent (FER-B) to enable cell lysis. Nucleic acid released during the lysis step hybridizes to magnetic particles in the wFCR-B. The capture particles are then separated from residual specimen matrix in a magnetic field by a series of wash steps with a mild detergent. The captured nucleic acid is then eluted from the magnetic particles with a reagent of low ionic strength (Panther Fusion Elution Buffer).

Note: The Panther Fusion System adds the IC-B to the Panther Fusion Capture Reagent-B (FCR-B). After the IC-B is added to the FCR-B, it is referred to as wFCR-B (working FCR-B).

Multiplex PCR amplification and fluorescence detection: Lyophilized single unit dose reaction master mix is reconstituted with the Panther Fusion Reconstitution Buffer I and then combined with the eluted nucleic acid into a reaction tube. Panther Fusion Oil reagent is added to prevent evaporation during the PCR reaction.

Target-specific primers and probes then amplify targets via polymerase chain reaction while simultaneously measuring fluorescence of the multiplexed targets. The Panther Fusion System compares the fluorescence signal to a predetermined cutoff to produce a qualitative result for the presence or absence of each analyte.

The analytes and the channel used for their detection on the Panther Fusion System are summarized in the table below:

Analyte	Gene Targeted	Instrument Channel
<i>Yersinia enterocolitica</i>	<i>InvA</i> (Invasive antigen A)	FAM
<i>Vibrio parahaemolyticus</i>	<i>gyrB</i> (Gyrase B)	HEX
<i>Vibrio vulnificus</i>	<i>gyrB</i> (Gyrase B)	HEX
<i>Vibrio cholerae</i>	<i>ompW</i> (Outer Membrane Protein W)	HEX
<i>Escherichia coli</i> O157	<i>rfbE</i> (Perosamine synthase-O-antigen)	ROX
<i>Plesiomonas shigelloides</i>	<i>hugA</i> (Heme utilization gene A)	RED647
Internal Control	Not Applicable	RED677

Warnings and Precautions

- A. For *in vitro* diagnostic use.
- B. Carefully read this entire package insert and the Panther®/Panther Fusion System Operator's Manual.
- C. The Panther Fusion Enhancer Reagent-B (FER-B) is corrosive, harmful if swallowed, and causes severe skin burns and eye damage.
- D. Only personnel adequately trained on the use of this assay and in handling potentially infectious materials should perform these procedures. If a spill occurs, immediately disinfect using appropriate site procedures.

Laboratory Related

- E. Use only supplied or specified disposable laboratory ware.
- F. Wear disposable, powderless gloves, protective eye wear, and laboratory coats when handling specimens and reagents. Wash hands thoroughly after handling specimens and reagents.
- G. Dispose of all material that has come into contact with specimens and reagents in accordance with applicable national, international, and regional regulations.

Specimen Related

- H. Handle all specimens as if infectious, using safe laboratory procedures such as those outlined in CDC/NIH Biosafety in Microbiological and Biomedical Laboratories and in the CLSI Document M29 Protection of Laboratory Workers from Occupationally Acquired Infections.⁹
- I. Expiration dates listed on the Aptima® Multitest tubes pertain to the transfer of sample into the tube and not to testing of the sample. Specimens collected/transferred any time prior to these expiration dates are valid for testing provided they are transported and stored in accordance with the appropriate package insert, even if these expiration dates have passed.
- J. Maintain proper storage conditions during specimen shipping to ensure the integrity of the specimen. Specimen stability under shipping conditions other than those recommended has not been evaluated.
- K. Avoid cross-contamination during the specimen handling steps. Specimens can contain extremely high levels of bacteria or other organisms. Ensure that specimen containers do not come in contact with one another, and discard used materials without passing them over any open containers. Change gloves if they come in contact with specimens.

Assay Related

- L. Do not use the reagents and controls after the expiration date.
- M. Store assay components at the recommended storage condition. See *Reagent Storage and Handling Requirements* and *Panther Fusion System Test Procedure* for more information.

- N. Do not combine any assay reagents or fluids. Do not top off reagents or fluids; the Panther Fusion System verifies reagent levels.
- O. Avoid microbial and nuclease contamination of reagents.
- P. Quality control requirements must be performed in conformance with local, state, and/or federal regulations or accreditation requirements and your laboratory's standard quality control procedures.
- Q. Do not use the assay cartridge if the storage pouch is compromised or if the assay cartridge foil is not intact. Contact Hologic® Technical Support if either occurs.
- R. Do not use the fluid packs if the foil seal is leaking. Contact Hologic Technical Support if this occurs.
- S. Handle the assay cartridges with care. Do not drop or invert assay cartridges. Avoid prolonged exposure to ambient light.
- T. Some reagents of this kit are labeled with hazard information.

Note: For more information on any hazard and precautionary statements that may be associated with reagents refer to the Safety Data Sheet Library at www.hologicsds.com. For more information on the symbols, refer to the symbol legend on <http://www.hologic.com/package-inserts>.

US Hazard Information**Panther Fusion Enhancer Reagent-B (FER-B)**

Lithium Hydroxide, Monohydrate 5 - 10%

**DANGER**

H302 - Harmful if swallowed

H314 - Causes severe skin burns and eye damage

P264 - Wash face, hands and any exposed skin thoroughly after handling

P270 - Do not eat, drink or smoke when using this product

P330 - Rinse mouth

P501 - Dispose of contents/ container to an approved waste disposal plant.

P260 - Do not breathe dusts or mists

P280 - Wear protective gloves/protective clothing/eye protection/face protection

P301 + P330 + P331 - IF SWALLOWED: rinse mouth. Do NOT induce vomiting

P303 + P361 + P353 - IF ON SKIN (or hair): Take off immediately all contaminated clothing. Rinse skin with water/shower

P304 + P340 - IF INHALED: Remove person to fresh air and keep comfortable for breathing

P305 + P351 + P338 - IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing

P321 - Specific treatment (see supplemental first aid instructions in the SDS)

P363 - Wash contaminated clothing before reuse

P405 - Store locked up

P301+ P317 - IF SWALLOWED: Get medical help

P316 - Get emergency medical help immediately

Reagent Storage and Handling Requirements

A. The following table provides storage and handling requirements for this assay.

Reagent	Unopened Storage	On Board/ Open Stability ^a	Opened Storage
Panther Fusion GI Expanded Bacterial Assay Cartridge	2°C to 8°C	60 days	2°C to 8°C ^b
Panther Fusion Capture Reagent-B (FCR-B)	15°C to 30°C	30 days	15°C to 30°C
Panther Fusion Enhancer Reagent-B (FER-B)	15°C to 30°C	30 days	15°C to 30°C
Panther Fusion Internal Control-B (IC-B)	2°C to 8°C	(In wFCR-B)	Not applicable
Panther Fusion Elution Buffer	15°C to 30°C	60 days	15°C to 30°C
Panther Fusion Oil	15°C to 30°C	60 days	15°C to 30°C
Panther Fusion Reconstitution Buffer I	15°C to 30°C	60 days	15°C to 30°C
Panther Fusion GI Expanded Bacterial Positive Control	2°C to 8°C	Single use vial	Not applicable-single use
Panther Fusion Negative Control	2°C to 8°C	Single use vial	Not applicable-single use

When reagents are removed from the Panther Fusion System, return them immediately to their appropriate storage temperatures.

^a On-board stability starts at the time the reagent is placed on the Panther Fusion System for the Panther Fusion GI Expanded Bacterial Assay cartridge, FCR-B, FER-B, and IC-B. On-board stability for the Panther Fusion Reconstitution Buffer I, Panther Fusion Elution Buffer, and Panther Fusion Oil Reagent starts when the reagent pack is first used.

^b If removed from the Panther Fusion System, store the assay cartridge in an air-tight container with desiccant at the recommended storage temperature.

- B. Working Panther Fusion Capture Reagent-B (wFCR-B) and Panther Fusion Enhancer Reagent-B (FER-B) are stable for 60 days when capped and stored at 15°C to 30°C. Do not refrigerate.
- C. Controls are stable until the date indicated on the vials.
- D. Discard any unused reagents that have surpassed their on board stability.
- E. Avoid cross-contamination during reagent handling and storage.
- F. **Do not freeze reagents.**

Specimen Collection and Storage

Specimens – Clinical material collected from patient and placed in an appropriate transport system. For the Panther Fusion GI Expanded Bacterial Assay, this includes raw stool preserved in Cary-Blair transport media.

Samples – Represents a more generic term to describe any material for testing on the Panther Fusion System including specimens, specimens transferred into an Aptima Multitest tube and controls.

Note: *Handle all specimens as if they contain potentially infectious agents. Use Universal Precautions.*

Note: *Take care to avoid cross-contamination during specimen handling steps. For example, discard used material without passing over open tubes.*

A. Specimen types include stool samples preserved in Cary-Blair transport media.

Collect raw stool following appropriate standard stool collection and handling procedures. Transfer raw stool specimens into Cary-Blair transport media according to manufacturer's instructions.

B. Specimen Processing

1. Mix Cary-Blair preserved specimen thoroughly to ensure homogeneity immediately prior to transfer into the Aptima Multitest tube.
2. Prior to testing on the Panther Fusion System, transfer specimen to an Aptima Multitest tube.

- a. Partially peel open the swab package. Remove the swab. Do not touch the soft tip or lay the swab down. If the soft tip is touched, the swab is laid down, or the swab is dropped, use a new Aptima® Multitest Swab Specimen Collection Kit. Completely submerge the soft tip of the swab in Cary-Blair preserved stool specimen.

Note: *Submerge only the soft tip of the swab 1 time in the liquid part, ensuring the pink shaft is not submerged.*

- b. Uncap the Aptima Multitest tube containing the transport medium. If the contents of the tube are spilled, use a new Aptima Multitest Swab Specimen Collection Kit. Place the swab in the tube and gently swirl the swab in the tube for 5 seconds to release material. Leave the swab in the tube.
- c. Carefully break the swab shaft at the score line against the side of the tube and discard the top portion of the swab shaft.
- d. Affix the provided or new penetrable cap to the tube.

3. Storing specimens before testing

- a. After collection, the Cary-Blair preserved specimens can be stored at 2°C to 8°C for up to 72 hours before transfer to the Aptima Multitest tube.

Note: *Yersinia is affected by storage temperature and time. If samples are not stored appropriately, they may have reduced recovery and lose their positive results.*

- b. Specimen in the Aptima Multitest tube may be stored under 1 of the following conditions:

- 15°C to 30°C for up to 6 days or
- 2°C to 8°C for up to 30 days or
- ≤ -20°C for up to 3 months

Note: Minimize freeze-thaw cycles to prevent potential sample degradation.

Note: It is recommended that specimens transferred to the Aptima Multitest tube are stored capped and upright in a rack.

C. Specimen Storage after Testing

1. Samples that have been assayed should be stored upright in the rack under 1 of the following conditions:

- 15°C to 30°C for up to 6 days or
- 2°C to 8°C for up to 30 days or
- ≤ -20°C for up to 3 months

Note: Minimize freeze-thaw cycles to prevent potential sample degradation.

2. The samples should be covered with a new, clean plastic film or foil barrier.
3. If assayed samples need to be frozen or shipped, remove the penetrable cap and place a new non-penetrable cap on the specimen tubes. If samples need to be shipped for testing at another facility, recommended temperatures must be maintained. Prior to uncapping, specimen transport tubes must be kept upright for 5 minutes to bring all of the liquid down to the bottom of the tube. Avoid splashing and cross-contamination. Do not centrifuge.

Specimen Transport

Maintain specimen storage conditions during transport as described under *Specimen Collection and Storage*.

Note: Specimens must be shipped in accordance with applicable national, international, and regional transportation regulations.

Panther Fusion System

The Panther Fusion System is an integrated nucleic acid testing System that fully automates all steps necessary to perform various Panther Fusion assays from sample processing through amplification, detection, and data reduction.

Reagents and Materials Provided for Panther Fusion GI Expanded Bacterial Assay

Assay Packaging

Components	Part No.	Storage
Panther Fusion GI Expanded Bacterial Assay Cartridge 96 Tests Panther Fusion GI Expanded Bacterial assay cartridge, 12 tests, 8 per box	PRD-07121	2°C to 8°C
Panther Fusion Internal Control-B 960 Tests Panther Fusion Internal Control-B tube, 4 per box	PRD-06234	2°C to 8°C
Panther Fusion GI Expanded Bacterial Assay Controls Panther Fusion GI Expanded Bacterial Positive Control tube, 5 per box Panther Fusion Negative Control tube, 5 per box	PRD-07122	2°C to 8°C
Panther Fusion Extraction Reagent-B 960 Tests Panther Fusion Capture Reagent-B bottle, 240 tests, 4 per box Panther Fusion Enhancer Reagent-B bottle, 240 tests, 4 per box	PRD-06232	15°C to 30°C
Panther Fusion Elution Buffer 2400 Tests Panther Fusion Elution Buffer pack, 1200 tests, 2 per box	PRD-04334	15°C to 30°C
Panther Fusion Reconstitution Buffer I 1920 Tests Panther Fusion Reconstitution Buffer I, 960 Tests, 2 per box	PRD-04333	15°C to 30°C
Panther Fusion Oil Reagent 1920 Tests Panther Fusion Oil Reagent, 960 tests, 2 per box	PRD-04335	15°C to 30°C

Individually Packaged Items

Items	Part No.
Panther Fusion Tube Trays, 1008 Tests, 18 trays per box	PRD-04000
Aptima Multitest Specimen Collection Kit, pack of 50	PRD-03546

Materials Required and Available Separately

Note: Materials available from Hologic have catalog numbers listed, unless otherwise specified.

Material	Cat. No.
Panther System	303095
Panther Fusion System	PRD-04172
Panther System Continuous Fluid and Waste (Panther Plus)	PRD-06067
Aptima® Assay Fluids Kit (Aptima Wash Solution, Aptima Buffer for Deactivation Fluid, and Aptima Oil Reagent)	303014 (1000 tests)
Multi-tube units (MTUs)	104772-02
Panther Waste Bag Kit	902731
Panther Waste Bin Cover	504405
Or Panther System Run Kit contains MTUs, waste bags, waste bin covers, assay fluids, and auto detects ^a	303096 (5000 tests)
Tips, 1000 µL, filtered, liquid-sensing, conductive, and disposable:	901121 (10612513 Tecan) 903031 (10612513 Tecan)
Not all products are available in all regions. Contact your representative for region-specific information.	MME-04134 (30180117 Tecan) MME-04128 MME-04110
Aptima penetrable caps (optional)	105668
Replacement non-penetrable caps (optional)	103036A
Replacement extraction reagent bottle caps	CL0040
Bleach, 5% to 8.25% (0.7 M to 1.16 M) sodium hypochlorite solution	
Note: Refer to the <i>Panther/Panther Fusion System Operator's Manual</i> for instructions on preparing diluted sodium hypochlorite solution.	—
Disposable powderless gloves	—

^a Needed only for Aptima assays that use TMA technology.

Optional Materials

Material	Cat. No.
Benchtop Vortex (VWR Analog Vortex Mixer 120V, Cat. No. 10153-838) or equivalent	—

Panther Fusion System Test Procedure

Note: Refer to the *Panther/Panther Fusion System Operator's Manual* for additional procedural information.

A. Work Area Preparation

1. Wipe down work surfaces with 2.5% to 3.5% (0.35 M to 0.5 M) sodium hypochlorite solution. Allow the sodium hypochlorite solution to contact surfaces for at least 1 minute and follow with a deionized (DI) water rinse. Do not allow the sodium hypochlorite solution to dry. Cover the bench surface with clean, plastic-backed absorbent laboratory bench covers.

B. Reagent Preparation

1. Remove the bottles of IC-B, FCR-B, and FER-B from storage.
2. Mix FCR-B by gently swirling until full resuspension of the beads. Avoid creating foam during this step.
3. Open the bottles of IC-B, FCR-B, and FER-B, and discard the caps. Open the TCR door on the upper bay of the Panther Fusion System.
4. Place the IC-B, FCR-B, and FER-B bottles in the appropriate positions on the TCR carousel.
5. Close the TCR door.

Note: The Panther Fusion System adds the IC-B to the FCR-B. After the IC-B is added to the FCR-B, it is referred to as wFCR-B (working FCR-B). If the wFCR-B and FER-B are removed from the System, use new caps and immediately store according to the proper storage conditions.

C. Specimen Handling

1. Visually confirm that each specimen tube contains a single pink Aptima collection swab in the Aptima Multitest tube. If the Aptima Multitest tube contains no swab, multiple swabs, or a swab not provided by Hologic, the transfer of stool in Cary-Blair media should be repeated using a new Aptima Multitest Swab Specimen Collection Kit.
2. Verify the appearance of the sample in the Aptima Multitest tube.
 - a. If the specimen is homogeneous, proceed with testing.
 - b. If solids or mucoidal materials are observed, note that these can interfere with the test.

Note: If any invalid flags are observed when processing specimens (e.g., CLT, icrfu, ebh or ebl), samples in the Aptima Multitest tube may be vortexed after replacing with a new penetrable cap for 30 to 60 seconds at maximum speed on a standard bench top vortex prior to retesting.

Note: Prepare specimens per the Specimen Processing instructions in the Specimen Collection and Storage section before loading specimens onto the Panther Fusion System.

D. System Preparation

For instructions on setting up the Panther Fusion System including loading samples, reagents, assay cartridges and universal fluids, refer to the *Panther/Panther Fusion System Operator's Manual*.

Procedural Notes

A. Controls

1. The Panther Fusion GI Expanded Bacterial Positive Control and the Panther Fusion Negative Control can be loaded in any rack position, in any Sample Bay lane on the Panther Fusion System.
2. Once the control tubes are pipetted and processed for the Panther Fusion GI Expanded Bacterial Assay, they are valid for up to 30 days (control frequency configured by an administrator) unless control results are invalid or a new assay cartridge lot is loaded.
3. Each control tube can be tested once.
4. Patient specimen pipetting begins when 1 of the following 2 conditions is met:
 - a. Valid results for the controls are registered on the system.
 - b. A pair of controls is currently in process on the system.

Quality Control

A run or specimen result may be invalidated by the Panther Fusion System if problems occur while performing the assay. Specimens with invalid results must be retested.

Negative and Positive Controls

To generate valid results, a set of assay controls must be tested. One (1) replicate of the negative assay control and positive assay control must be tested each time a new lot of assay cartridges is loaded on the Panther Fusion System or when the current set of valid controls for an active cartridge lot have expired.

The Panther Fusion System is configured to require assay controls run at an administrator-specified interval of up to 30 days. Software on the Panther Fusion System alerts the operator when assay controls are required and does not start new tests until the assay controls are loaded and have started processing.

During processing, criteria for acceptance of the assay controls are automatically verified by the Panther Fusion System. To generate valid results, the assay controls must pass a series of validity checks performed by the Panther Fusion System.

If the assay controls pass all validity checks, they are considered valid for the administrator-specified time interval. When the time interval has passed, the assay controls are expired by the Panther Fusion System and a new set of assay controls will be required prior to starting any new samples.

If any one of the assay controls fails the validity checks, the Panther Fusion System automatically invalidates the affected samples and a new set of assay controls will be required prior to testing any new samples.

Internal Control

An internal control is added to each sample during the extraction process. During processing, the internal control acceptance criteria is automatically verified by the Panther Fusion System software. Detection of the internal control is not required for samples that are positive for *Yersinia enterocolitica*, *Vibrio* species, *Escherichia coli* O157, and/or *Plesiomonas shigelloides*. The internal control must be detected in all samples that are negative for all of the intended analytes; samples that fail to meet that criteria will be reported as Invalid. Each sample with an Invalid result must be retested.

The Panther Fusion System is designed to accurately verify processes when procedures are performed following the instructions provided in this package insert and the *Panther/Panther Fusion System Operator's Manual*.

Interpretation of Results

The Panther Fusion System automatically determines the test results for samples and controls. Results for *Yersinia enterocolitica*, *Vibrio* species, *Escherichia coli* O157, and *Plesiomonas shigelloides* detection are reported separately. A test result may be negative, positive, or invalid.

The first valid result is the result that should be reported. Samples with invalid results should be retested. If the result is invalid upon retest, a new specimen should be collected.

Table 1 shows the possible results reported in a valid run with corresponding result interpretations.

Table 1: Result Interpretation

Yersinia Result	Vibrio Result	O157 Result ^a	Plesio Result	IC Result	Interpretation
Neg	Neg	Neg	Neg	Valid	<i>Yersinia enterocolitica</i> , <i>Vibrio</i> species, <i>E. coli</i> O157, and <i>Plesiomonas shigelloides</i> not detected.
POS	Neg	Neg	Neg	Valid	<i>Yersinia enterocolitica</i> detected.
Neg	POS	Neg	Neg	Valid	<i>Vibrio</i> species detected.
Neg	Neg	POS	Neg	Valid	<i>E. coli</i> O157 detected.
Neg	Neg	Neg	POS	Valid	<i>Plesiomonas shigelloides</i> detected.
POS	POS	Neg	Neg	Valid	<i>Yersinia enterocolitica</i> and <i>Vibrio</i> species detected.
POS	Neg	POS	Neg	Valid	<i>Yersinia enterocolitica</i> and <i>E. coli</i> O157 detected.
POS	Neg	Neg	POS	Valid	<i>Yersinia enterocolitica</i> and <i>Plesiomonas shigelloides</i> detected.
Neg	POS	POS	Neg	Valid	<i>Vibrio</i> species and <i>E. coli</i> O157 detected.
Neg	POS	Neg	POS	Valid	<i>Vibrio</i> species and <i>Plesiomonas shigelloides</i> detected.
Neg	Neg	POS	POS	Valid	<i>E. coli</i> O157 and <i>Plesiomonas shigelloides</i> detected.
POS	POS	POS	Neg	Valid	<i>Yersinia enterocolitica</i> , <i>Vibrio</i> species, and <i>E. coli</i> O157 detected. Infections with 3 bacteria are rare. Retest to confirm result.
POS	POS	Neg	POS	Valid	<i>Yersinia enterocolitica</i> , <i>Vibrio</i> species, and <i>Plesiomonas shigelloides</i> detected. Infections with 3 bacteria are rare. Retest to confirm result.
POS	Neg	POS	POS	Valid	<i>Yersinia enterocolitica</i> , <i>E. coli</i> O157, and <i>Plesiomonas shigelloides</i> detected. Infections with 3 bacteria are rare. Retest to confirm result.
Neg	POS	POS	POS	Valid	<i>Vibrio</i> species, <i>E. coli</i> O157, and <i>Plesiomonas shigelloides</i> detected. Infections with 3 bacteria are rare. Retest to confirm result.
POS	POS	POS	POS	Valid	<i>Yersinia enterocolitica</i> , <i>Vibrio</i> species, <i>E. coli</i> O157, and <i>Plesiomonas shigelloides</i> detected. Infections with 4 bacteria are rare. Retest to confirm result.
Invalid	Invalid	Invalid	Invalid	Invalid	Invalid. There was an error in the generation of the result; retest specimen.

Neg = negative, POS = positive.

Note: POS result will be accompanied by cycle threshold (Ct) values. POS/HT represents a high titer result and will not have a Ct reported.

^a The Panther Fusion GI Bacterial Assay provides results for Shiga-toxin genes *stx1/stx2*. Note that strains of *E. coli* O157 that do not carry the Shiga-like toxin genes have been identified. However, the clinical significance of these non-STEC O157 strains has not been established.

Limitations

- A. Use of this assay is limited to personnel who are trained in this procedure. Failure to follow these instructions may result in erroneous results.
- B. Reliable results are dependent on adequate specimen collection, transport, storage, and processing.
- C. Avoid contamination by adhering to good laboratory practices and to the procedures specified in this package insert.
- D. Dehydrated Cary-Blair medium powders and Cary-Blair media in solid configuration with high agarose content were not evaluated and may not be compatible with the assay sample processing steps.
- E. The performance of this test has only been validated with human stool preserved in liquid Cary-Blair transport medium, according to the media manufacturers' instructions. It has not been validated for use with other stool transport media, raw stool, rectal swabs, fecal swabs, endoscopy stool aspirates, or vomitus.
- F. *Vibrio* at concentrations higher than 10^4 CFU/mL may compete with *Yersinia enterocolitica* and STEC O157 at low concentrations near LoD resulting in false negative results for these 2 pathogens.
- G. The presence of *Vibrio alginolyticus* at concentrations higher than 10^4 CFU/mL may cross react with the detection of *Vibrio* (*V. cholera*, *V. parahaemolyticus*, *V. vulnificus*) resulting in false positive results.
- H. Cross-reactivity with organisms other than those listed in the *Analytical Specificity* section have not been evaluated.
- I. The effect of interfering substances has only been evaluated for those listed in this package insert. Potential interference has not been evaluated for substances other than those described in the *Interference* section.
- J. This product should not be used to test stool samples in fixative, for example formalin and low viscosity polyvinyl alcohol (LV-PVA).

Analytical Performance

Analytical Sensitivity

The analytical sensitivity (Limit of Detection or LoD) of the Panther Fusion GI Expanded Bacterial Assay was determined by testing dilutions of preserved negative Cary-Blair stool processed with STM (negative CBS matrix) spiked with bacterial cultures of *Yersinia* (2 strains), *Vibrio* (3 strains), STEC O157 (2 strains), and *Plesiomonas* (2 strains). A minimum of 24 replicates were tested with each of the 3 reagent lots. The LoD for each analyte was determined by Probit analysis for each reagent lot and was confirmed with an additional 24 replicates using a single reagent lot in single analyte and multi-analyte configuration. Analytical sensitivity is defined as the lowest concentration at which $\geq 95\%$ of all replicates tested positive, as summarized in Table 2.

Table 2: Analytical Sensitivity

Strain	LoD Concentration (CFU/mL) ^a	
	Aptima Multitest Tube	Preserved Stool
<i>Yersinia enterocolitica</i> , 33114	91	1,820
<i>Yersinia enterocolitica</i> , 1375, O:8	94	1,880
<i>Vibrio parahaemolyticus</i> , EB101	90	1,800
<i>Vibrio vulnificus</i> , B9629	10	200
<i>Vibrio cholerae</i> , 8021	33	660
STEC O157:H7, EDL 931	53	1,060
O157:NM, CDC 92-3073	357	7,140
<i>Plesiomonas shigelloides</i> , CDC 3085-55	65	1,300
<i>Plesiomonas shigelloides</i> , GNI 14	34	680

CFU = colony forming units.

^a Analyte concentrations in Aptima Multitest tube are ~20X dilute compared to preserved stool (~150 μ L preserved stool in ~3 mL)

Inclusivity/Reactivity - Wet Testing

The inclusivity/reactivity of the Panther Fusion GI Expanded Bacterial Assay was determined by testing bacterial strains in negative CBS matrix. Each strain was tested in triplicate at 3X LoD with 1 reagent lot in single or multi-analyte configuration. For strains not detected at 3X LoD, additional testing at higher concentrations was performed until 100% positivity was observed. Table 3 shows the lowest concentration of each strain at which 100% positivity was observed.

Table 3: Inclusivity/Reactivity Summary for the GI Expanded Bacterial Assay Analytes

Organism	ATCC# or Source	Strain/ Serovar/ Serotype/ Antigenic properties	Test Concentration (3X LoD) (CFU/mL)	
			Aptima Multitest Tube	Preserved Stool
<i>Yersinia enterocolitica</i>	BEI NR-207	CDC 497-70, O:8	282	5,640
	BEI NR-212	NCTC 11175, O:3	282	5,640
	23715	Billups-1803-68, O:8	282	5,640
	49397	1375, O:8 ^c	282	5,640
	NCTC 10463	P 77, O:5, 27	282	5,640
	CCUG 4588	Type 2, O:9	282	5,640
	CCUG 8050	N/A	282	5,640
	CCUG 8232	Type 5, O:1, 2, 3 O:2, 3 O:3/XI	282	5,640
	CCUG 8234	Type 4	282	5,640
	55075	O:9	282	5,640
	27729	WA, Type 1, O:8	282	5,640
<i>Vibrio parahaemolyticus</i>	BEI NR-21990	48057, O4: K12	270	5,400
	BEI NR-21992	KXV 755, O4: K41	270	5,400
	BAA-242	VP250, O1:KUT	270	5,400
	27969	FC 1011	270	5,400
	BAA-241	VP232, O4:K68	270	5,400
	33845	117 [CDC KC830]	270	5,400
	43996	NCTC 10884 [70/116655]	270	5,400
	33846	205 [9302]	270	5,400
	49529	MDL 3875-7-83, O4:K12	270	5,400
	CCUG 34902	N/A	270	5,400
	CCUG 67711	N/A	270	5,400
	33847	279 [11590]	270	5,400
<i>Vibrio vulnificus</i>	33817	329 [CDC B3547], Biotype 2	33	660
	BAA-86	CDC 9505-95	33	660
	CCUG 38297	N/A	33	660
	CCUG 47321	N/A	33	660
	29306	CDC A1402 [P. Baumann 328]	33	660
	43382	VVL1	33	660
	29307	CDC A8694	33	660
	CCUG 38297	N/A ^b	55	1,110

Table 3: Inclusivity/Reactivity Summary for the GI Expanded Bacterial Assay Analytes (continued)

Organism	ATCC# or Source	Strain/ Serovar/ Serotype/ Antigenic properties	Test Concentration (3X LoD) (CFU/mL)	
			Aptima Multitest Tube	Preserved Stool
<i>Vibrio cholerae</i>	BEI NR-147	N16961, O:1	99	1,980
	BEI NR-148	CVD 101, O:1	99	1,980
	BEI NR-149	Nanking 32/123, O:2	99	1,980
	BEI NR-152	Nanking 32/124 (NCTC 8042), O:7	99	1,980
	14033	NCTC 8457 [R. Hugh 1092], O1, Inaba	99	1,980
	9459	AMC 20-A-10 [R. Hugh 583], Inaba	99	1,980
	CCUG 2573	NAG/NCV	99	1,980
	CCUG 2569	NAG/NCV	99	1,980
	CCUG 4070	Non O-1	99	1,980
	CCUG 21589	18	99	1,980
	CCUG 56875	N/A	99	1,980
	CCUG 53725	O1/O139	99	1,980
	CCUG14542	NA	99	1,980
	9458	AMC 20-A-41 [R. Hugh 582], Ogawa	99	1,980
	25870	569B	99	1,980
STEC O157: H7	43890	CDC C984 [CDC 3526-87], H7	159	3,180
	43895	CDC EDL 933, H7	159	3,180
	43894	CDC EDL 932, H7	159	3,180
	700927	EDL 933, H7:K-	159	3,180
STEC O157: NM	700375	CDC 94-G7771, NM	1,197	23,940
	700377	CDC 92-3099, NM	1,197	23,940
	700378	CDC 92-3073, NM ^c	1,197	23,940
	AR Bank # 427 ^a	N/A	1,197	23,940
	AR Bank # 428 ^a	N/A	1,197	23,940
	AR Bank # 429 ^a	N/A	1,197	23,940
	AR Bank # 430 ^a	N/A	1,197	23,940
<i>Plesiomonas shigelloides</i>	14030	CDC 16408 [Ferguson and Henderson C27, RH 864], O:17	195	3,900
	51903	GNI 14 ^c	195	3,900
	51572	CIP 69.35 [2886]	195	3,900
	CCUG 7041A	O17: H2	195	3,900
	CCUG 9221	O17	195	3,900
	CCUG 14309	O17: H2	195	3,900
	CCUG 14597	N/A	195	3,900

CFU = colony forming units.

^a These strains were evaluated using the higher LoD of the 2 serotypes which is the NM serotype.^b For this strain 100% positivity was observed at ~5X LoD. *In silico* analysis showed 100% homology to the amplification region.^c Strains used to establish LoD.

Inclusivity/Reactivity - *In Silico* Analysis

The inclusivity of the Panther Fusion GI Expanded Bacterial Assay was evaluated using *in silico* inclusivity analysis for each analyte. *In silico* analysis was performed using analyte sequences available in the NCBI database and in the whole genome shotgun sequence database. For each analyte, corresponding oligonucleotide sequences (primers and probes) were evaluated against the database sequences. Any sequences with insufficient lengths (not covering the entire amplicon region) were excluded from the analysis.

Based on *in silico* analysis of all sequences available up to May 30, 2023 in the databases, the Panther Fusion GI Expanded Bacterial Assay is predicted to detect 99.9% of 1,054 *Yersinia Enterocolitica*, 99.5% of 1,337 *Vibrio parahaemolyticus*, 99.1% of 1,180 *Vibrio vulnificus*, 98.0% of 1,189 *Vibrio cholerae*, 100% of 2,004 STEC O157, and 91.5% of 47 *Plesiomonas shigelloides* sequences evaluated.

Analytical Specificity: Cross Reactivity and Microbial Interference - Wet Testing

Analytical specificity (cross-reactivity) and microbial interference for the Panther Fusion GI Expanded Bacterial Assay were evaluated in the presence of non-targeted microorganisms that are either phylogenetically related to the assay analytes or potentially found in clinical specimens. Panels consisting of 109 bacteria, viruses, parasites, and yeast listed in Table 4 were tested in negative CBS matrix in the absence and in the presence of Panther Fusion GI Expanded Bacterial Assay analytes at 3X LoD. Except where noted, bacteria, yeast, and parasites were evaluated at 10^6 CFU/mL or 10^6 rRNA copies/mL or 10^6 cells/mL; viruses were evaluated at 10^5 TCID₅₀/mL. If cross-reactivity or interference was observed in the initial testing, then the organism was tested at lower concentrations until the expected result was observed. No cross-reactivity or microbial interference was observed with any of the organisms tested on the Panther Fusion GI Expanded Bacterial Assay at the indicated concentrations.

Table 4: Microorganisms tested for Cross-Reactivity and Microbial Interference

Microorganism	Test Concentration	Microorganism	Test Concentration
<i>Arcobacter cryaerophilus</i>	10^6 CFU/mL	<i>Enterococcus faecalis</i>	10^6 CFU/mL
<i>Neisseria gonorrhoeae</i>	10^6 CFU/mL	<i>Enterobacter aerogenes</i>	10^6 CFU/mL
<i>Streptococcus pyogenes</i>	10^6 CFU/mL	<i>Enterobacter cloacae</i>	10^6 CFU/mL
<i>Trabulsiella guamensis</i>	10^6 CFU/mL	<i>Escherichia fergusonii</i>	10^6 CFU/mL
<i>Faecalibacterium prausnitzii</i>	10^6 rRNA copies /mL	<i>Escherichia hermannii</i>	10^6 CFU/mL
<i>Escherichia coli</i> (non-shigatoxigenic)	10^6 CFU/mL	<i>Escherichia vulneris</i>	10^6 CFU/mL
<i>Giardia lamblia</i> BG-A ^a	10^6 copies/mL	<i>Gardnerella vaginalis</i>	10^6 CFU/mL
<i>Cyclospora</i> ^a	10^6 copies/mL	<i>Helicobacter pylori</i>	10^6 CFU/mL
<i>Cryptosporidium</i> ^a	10^6 copies/mL	<i>Klebsiella oxytoca</i>	10^6 CFU/mL
Norovirus (Noro GII) ^a	10^6 copies/mL	<i>Klebsiella ozaenae</i>	10^6 CFU/mL
Astrovirus ^a	10^6 copies/mL	<i>Klebsiella pneumoniae</i>	10^6 CFU/mL
Sapovirus (GII) ^a	10^5 copies/mL	<i>Lactobacillus acidophilus</i>	10^6 CFU/mL
Enterovirus (Ent V) ^a	10^5 copies/mL	<i>Lactobacillus crispatus</i>	10^6 CFU/mL

Table 4: Microorganisms tested for Cross-Reactivity and Microbial Interference (continued)

Microorganism	Test Concentration	Microorganism	Test Concentration
Rhinovirus ^a	10 ⁵ copies/mL	<i>Lactococcus lactis</i>	10 ⁶ CFU/mL
Coronavirus 229E	10 ⁵ TCID ₅₀ /mL	<i>Listeria grayi</i>	10 ⁶ CFU/mL
Coxsackievirus Type B4	10 ⁵ TCID ₅₀ /mL	<i>Listeria monocytogenes</i>	10 ⁶ CFU/mL
Adenovirus Type 7A	10 ⁵ TCID ₅₀ /mL	<i>Morganella morganii</i>	10 ⁶ CFU/mL
Rotavirus ^a	10 ⁵ copies/mL	<i>Peptostreptococcus anaerobius</i>	10 ⁶ CFU/mL
<i>Anaerococcus tetradius</i>	10 ⁶ CFU/mL	<i>Peptostreptococcus micros</i>	10 ⁶ rRNA copies /mL
<i>Abiotrophia defectiva</i>	10 ⁶ CFU/mL	<i>Photobacterium damsela</i>	10 ⁶ CFU/mL
<i>Acinetobacter baumannii</i>	10 ⁶ CFU/mL	<i>Prevotella bivia</i>	10 ⁶ CFU/mL
<i>Acinetobacter lwoffii</i>	10 ⁶ CFU/mL	<i>Prevotella melaninogenica</i>	10 ⁶ CFU/mL
<i>Aeromonas hydrophila</i>	10 ⁶ CFU/mL	<i>Proteus mirabilis</i>	10 ⁶ rRNA copies /mL
<i>Alcaligenes faecalis</i>	10 ⁶ CFU/mL	<i>Proteus penneri</i>	10 ⁶ CFU/mL
<i>Campylobacter upsaliensis</i>	10 ⁶ CFU/mL	<i>Proteus vulgaris</i>	10 ⁶ CFU/mL
<i>Anaerococcus vaginalis</i>	10 ⁶ CFU/mL	<i>Providencia alcalifaciens</i>	10 ⁶ CFU/mL
<i>Arcobacter butzleri</i>	10 ⁶ CFU/mL	<i>Providencia rettgeri</i>	10 ⁶ CFU/mL
<i>Bacillus cereus</i>	10 ⁶ CFU/mL	<i>Providencia stuartii</i>	10 ⁶ CFU/mL
<i>Bacteriodes fragilis</i>	10 ⁶ CFU/mL	<i>Pseudomonas aeruginosa</i>	10 ⁶ CFU/mL
<i>Bacteroides thetaiotaomicron</i>	10 ⁶ CFU/mL	<i>Pseudomonas fluorescens</i>	10 ⁶ CFU/mL
<i>Bacteroides vulgatus</i>	10 ⁶ CFU/mL	<i>Serratia liquefaciens</i>	10 ⁶ CFU/mL
<i>Bifidobacterium adolescentis</i>	10 ⁶ CFU/mL	<i>Serratia marcescens</i>	10 ⁶ CFU/mL
<i>Bifidobacterium longum</i>	10 ⁶ rRNA copies /mL	<i>Staphylococcus aureus</i>	10 ⁶ CFU/mL
<i>Campylobacter fetus</i>	10 ⁶ CFU/mL	<i>Staphylococcus epidermidis</i>	10 ⁶ CFU/mL
<i>Campylobacter hyointestinalis</i>	10 ⁶ CFU/mL	<i>Stenotrophomonas maltophilia</i>	10 ⁶ CFU/mL
<i>Campylobacter rectus</i>	10 ⁶ CFU/mL	<i>Streptococcus anginosus</i>	10 ⁶ CFU/mL
<i>Campylobacter sputorum</i>	10 ⁶ CFU/mL	<i>Streptococcus dysgalactiae</i>	10 ⁶ CFU/mL
<i>Candida albicans</i>	10 ⁶ CFU/mL	<i>Yersinia bercovieri</i>	10 ⁶ CFU/mL
<i>Citrobacter freundii</i>	10 ⁶ CFU/mL	<i>Yersinia pseudotuberculosis</i>	10 ⁶ CFU/mL
<i>Citrobacter koseri</i>	10 ⁶ CFU/mL	<i>Yersinia rohdei</i>	10 ⁶ CFU/mL
<i>Clostridium difficile</i>	10 ⁶ CFU/mL	<i>Campylobacter lari</i>	10 ⁶ CFU/mL
<i>Clostridium perfringens</i>	10 ⁶ CFU/mL	<i>Entamoeba histolytica</i>	10 ⁴ cells/mL
<i>Clostridium ramosum</i>	10 ⁶ CFU/mL	<i>Megasphaera elsdenii</i>	10 ⁶ CFU/mL
<i>Clostridium sordellii</i>	10 ⁶ CFU/mL	<i>Chlamydia trachomatis</i>	10 ⁵ IFU/mL
<i>Clostridium tertium</i>	10 ⁶ CFU/mL	<i>Leptotrichia buccalis</i>	10 ⁶ CFU/mL
<i>Collinsella aerofaciens</i>	10 ⁶ CFU/mL	Cytomegalovirus	10 ⁵ TCID ₅₀ /mL
<i>Corynebacterium genitalium</i>	10 ⁶ CFU/mL	<i>Salmonella enterica</i>	10 ⁶ CFU/mL
<i>Cronobacter sakazakii</i>	10 ⁶ CFU/mL	<i>Campylobacter jejuni</i>	10 ⁶ CFU/mL
<i>Edwardsiella tarda</i>	10 ⁶ CFU/mL	<i>Shigella sonnei</i>	10 ⁶ CFU/mL
<i>Eggerthella lenta</i>	10 ⁶ rRNA copies /mL	STEC - stx1	10 ⁶ CFU/mL

Table 4: Microorganisms tested for Cross-Reactivity and Microbial Interference (continued)

Microorganism	Test Concentration	Microorganism	Test Concentration
STEC - <i>stx2</i>	10 ⁶ CFU/mL	<i>Vibrio mimicus</i>	10 ⁶ CFU/mL
<i>Vibrio fluvialis</i>	10 ⁶ CFU/mL	<i>Yersinia frederiksenii</i>	10 ⁶ CFU/mL
<i>Vibrio furnissii</i>	10 ⁶ CFU/mL	<i>Yersinia kristensenii</i>	10 ⁶ CFU/mL
<i>Vibrio metschnikovii</i>	10 ⁶ CFU/mL	<i>Vibrio alginolyticus</i> ^b	10 ⁴ CFU/mL

CFU = colony forming units, IFU = inclusion forming units, rRNA copies = ribosomal ribonucleic acid copies, TCID₅₀ = Median Tissue Culture Infectious Dose.

^a *In vitro* transcripts were used to evaluate cross-reactivity and microbial interference as cultured virus or whole genome purified nucleic acid are not readily available.

^b Cross reactivity was observed at concentrations ≥10⁵ CFU/mL.

Coinfection/Competitive Interference

Competitive interference in the Panther Fusion GI Expanded Bacterial Assay was evaluated in triplicate using pairs of assay analytes at low/high concentrations in negative CBS matrix. The low concentration analyte was tested at 3X LoD against a high concentration analyte at 10⁶ CFU/mL. Additionally, analytes were also tested in the absence of a second analyte. If less than 100% positivity was observed for the low concentration analyte, the high concentration analyte was diluted until a concentration was reached where 100% positivity was achieved for the low concentration analyte. The highest concentration of competing analyte at which the low concentration analyte maintained a 100% positivity is shown in Table 5. When the analytes were tested at high concentration, all results for other analytes maintained expected positivity; no competitive interference was observed.

Table 5: Summary of Coinfection Results

Analyte 1		Analyte 2		Yersinia % Pos	Vibrio % Pos	STEC O157 % Pos	Plesiomonas % Pos
Name	3X LoD (CFU/mL) ^a	Name	High Conc (CFU/mL) ^a				
Negative	NA	Negative	NA	0%	0%	0%	0%
Yersinia	282	None	0	100%	0%	0%	0%
		Vibrio ^b	10 ⁴	100%	100%	0%	0%
		STEC O157	10 ⁶	100%	0%	100%	0%
		Plesiomonas	10 ⁶	100%	0%	0%	100%
Vibrio	282	None	0	0%	100%	0%	0%
		Yersinia	10 ⁶	100%	100%	0%	0%
		STEC O157	10 ⁶	0%	100%	100%	0%
		Plesiomonas	10 ⁶	0%	100%	0%	100%

Table 5: Summary of Coinfection Results (continued)

STEC O157	1,197	None	0	0%	0%	100%	0%
		<i>Yersinia</i>	10 ⁶	100%	0%	100%	0%
		<i>Vibrio</i> ^b	10 ⁴	0%	100%	100%	0%
		<i>Plesiomonas</i>	10 ⁶	0%	0%	100%	100%
<i>Plesiomonas</i>	195	None	0	0%	0%	0%	100%
		<i>Yersinia</i>	10 ⁶	100%	0%	0%	100%
		<i>Vibrio</i> ^b	10 ⁶	0%	100%	0%	100%
		STEC O157	10 ⁶	0%	0%	100%	100%
None	0	<i>Yersinia</i>	10 ⁶	100%	0%	0%	0%
		<i>Vibrio</i>	10 ⁶	0%	100%	0%	0%
		STEC O157	10 ⁶	0%	0%	100%	0%
		<i>Plesiomonas</i>	10 ⁶	0%	0%	0%	100%

CFU = colony forming units, Conc = concentration, Pos = positive.

^a Analyte concentration in Aptima Multitest tube..

^b Less than 100% positive results were observed for analyte 1 with *Vibrio* at $\geq 10^5$ CFU/mL.

Interference

Potential inhibitory effects of endogenous and exogenous substances that may be present in a specimen were evaluated in the Panther Fusion GI Expanded Bacterial Assay. Clinically relevant concentrations of potentially interfering substances were added to negative CBS matrix and tested in the absence and in the presence of GI Expanded Bacterial Assay analytes at 3X LoD. Tests were performed in triplicate. The substances and test concentrations are shown in Table 6.

No impact on the performance of the Panther Fusion GI Expanded Bacterial Assay was observed for any of the substances at the concentrations tested.

Table 6: Substances Tested for Interference

Substance Type	Generic Name	Active Ingredient(s)	Test Concentration ^{a, b, c}
Antibiotics	Amoxicillin	Amoxicillin	0.7 µg/mL
	Ampicillin	Ampicillin	0.9 µg/mL
	Doxycycline	Doxycycline	0.2 µg/mL
	Metronidazole	Metronidazole	1.5 µg/mL
	Neosporin®	Polymyxin B sulfate, bacitracin zinc, neomycin sulfate	1.3% w/v
Antimicrobial and antifungal	BZK Antiseptic Towelettes	Benzalkonium chloride	1.3% v/v
	Nystatin	Nystatin	1.3% v/v
Laxatives and stool softeners	Dulcolax® suppository	Bisacodyl	75 ng/mL
	Colace®	Docusate sodium	3.0 µg/mL
	Fleet® mineral oil enema	Mineral oil	1.3% v/v
	Ex-Lax®	Sennosides	0.8 µg/mL
	Miralax®	Polyethylene glycol 3350	0.1 mg/mL
	Milk of Magnesia	Magnesium hydroxide, Aluminum hydroxide	1.3% v/v
	Visicol®	Sodium phosphate	53 ng/mL
Anti-diarrheal	Imodium	Loperamide hydrochloride	0.1 µg/mL
Anti-itch	Vagisil®	Benzocaine	1.3% w/v
	Preparation H®	Hydrocortisone	1.3% w/v
Anti-inflammatory	Phenylephrine hydrochloride (for hemorrhoids)	Phenylephrine hydrochloride	0.4 ng/mL
	Mesalazine (Rx only, for Crohn's disease/ ulcerative colitis)	Salicylic acid	0.4 µg/mL
	Aleve®	Naproxen sodium	4.5 µg/mL
Antacid	Pepto-Bismol®	Bismuth subsalicylate	1.3% v/v
	Tums®	Calcium carbonate	55 µg/mL
Radiopaque contrast material	Barium sulfate	Barium sulfate	0.1 mg/mL
Lubricants and skin protectants	K-Y® Personal Lubricant Jelly	Glycerin	1.3% w/v
	Vaseline® Original 100% Pure Petroleum Jelly White	Petrolatum	1.3% w/v
	Desitin®	Zinc oxide	1.3% w/v
Spermicide	Options Conceptrol® Vaginal Contraceptive Gel	Nonoxynol-9	1.3% w/v

Table 6: Substances Tested for Interference (continued)

Substance Type	Generic Name	Active Ingredient(s)	Test Concentration ^{a, b, c}
Endogenous	Cholesterol	Cholesterol	50 µg/mL
	Fatty acids	Palmitic acid	16 µg/mL
	Fatty acids	Stearic acid	34 µg/mL
	Triglycerides, total (Fecal fat, Intralipid)	Triglycerides	1.3% v/v
	Human bile	Bilirubin, conjugated	5.0 µg/mL
	Urine	Human urine	1.3% v/v
	Human whole blood	Blood/hemoglobin	1.3% v/v
	Mucin	Purified mucin protein	0.05% w/v

^a Substance concentration in Aptima Multitest tube.^b v/v: volume by volume.^c w/v: weight by volume.

Stool specimens prepared in various preservative media were evaluated for potential impact on the Panther Fusion GI Expanded Bacterial Assay performance. The preservative media evaluated include 7 different types of Cary-Blair transport media from different vendors and preservative media containing fixatives shown in Table 7. All media were tested with GI Expanded Bacterial Assay analytes at 3X LoD. Comparable performance was seen with all Cary-Blair media. Interference was observed when specimens were processed in media containing fixatives.

Table 7: Stool Preservative Media Tested for Interference

Cary-Blair Media
Culture & Sensitivity (C&S) Medium
Cary-Blair Transport Medium w/ Indicator
Para-Pak® C&S
Para-Pak® Enteric Plus
Cardinal Health™ C&S Stool Transport Vial
Protocol Cary-Blair Medium
Enteric Transport Media (ETM)
Fixative Media (interference was observed)
Fisher® 10% Buffered Formalin
Para-Pak® 10% Buffered Formalin
Para-Pak® LV-PVA

Carryover Contamination

Panther Fusion GI Bacterial Assay and GI Expanded Bacterial Assay belong to the same family of assays that both utilize Cary-Blair Stool as the sample type and follow identical assay processing steps. Carryover contamination was evaluated using Panther Fusion GI Bacterial Assay as a representative assay and demonstrated a 0% carryover rate.

Within Laboratory Precision/Repeatability

Panther Fusion GI Expanded Bacterial Assay within laboratory precision was evaluated with a 5-member panel consisting of assay analytes in negative CBS matrix. The 5-member panel included 1 negative, 2 single analyte (*Yersinia*), and 2 multi-analyte (with *Vibrio*, STEC O157, and *Plesiomonas*) panel members. The panels were tested by 3 operators on 2 runs per day, using 3 reagent lots on 3 Panther Fusion Systems over 9 days.

The panel members are described in Table 8, along with a summary of the agreement with the expected results, mean Ct, variability analysis between reagent lots, operators, instruments, days, between and within runs, and overall (total).

Table 8: Ct Variability Analysis Summary

Panel	Description	Analyte	Agreed/N	Agreement % ^a	Mean Ct	Between Lots		Between Instruments		Between Operators		Between Days		Between Runs		Within Run		Total	
						SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)
1	Negative	Negative (Internal Control)	162/162	100	28.0	0.11	0.39	0.32	1.15	0.00	0.00	0.00	0.00	0.12	0.42	0.14	0.51	0.39	1.39
2	Low Pos (1.5X LoD)	<i>Yersinia</i>	162/162	100	34.6	0.07	0.20	0.08	0.23	0.04	0.12	0.00	0.00	0.00	0.00	0.48	1.39	0.50	1.43
3	Mod Pos (3X LoD)	<i>Yersinia</i>	162/162	100	33.7	0.03	0.08	0.09	0.26	0.00	0.00	0.00	0.00	0.00	0.00	0.41	1.23	0.42	1.26
4	Low Pos (1.5X LoD)	<i>Vibrio</i>	162/162	100	33.7	0.12	0.35	0.07	0.21	0.01	0.04	0.00	0.00	0.17	0.52	0.23	0.69	0.32	0.95
		STEC O157	162/162	100	32.4	0.02	0.08	0.04	0.13	0.00	0.00	0.00	0.00	0.11	0.34	0.28	0.87	0.31	0.95
		<i>Plesiomonas</i>	162/162	100	33.8	0.08	0.25	0.05	0.14	0.00	0.00	0.00	0.00	<0.01	0.03	0.25	0.73	0.26	0.78
5	Mod Pos (3X LoD)	<i>Vibrio</i>	162/162	100	32.7	0.07	0.21	0.12	0.37	0.00	0.00	0.00	0.00	0.19	0.57	0.20	0.06	0.30	0.93
		STEC O157	162/162	100	31.3	0.02	0.08	0.06	0.20	0.00	0.00	0.03	0.10	0.00	0.00	0.21	0.68	0.22	0.72
		<i>Plesiomonas</i>	162/162	100	33.1	0.05	0.17	<0.01	0.03	0.01	0.03	0.06	0.17	0.00	0.00	0.19	0.56	0.20	0.61

Ct = cycle threshold, CV = coefficient of variation, Mod = moderate, N = sample size, Pos = positive, SD = standard deviation.

^a Agreement to expected panel positivity result.

Reproducibility

Panther Fusion GI Expanded Bacterial Assay reproducibility was evaluated at 3 US sites using 1 negative panel member and 4 panel members positive for 1 or 3 targets. Testing was performed for 5 days by 6 operators (2 at each site) using 1 lot of assay reagents. Each run included 3 replicates of each panel member.

A negative panel member was created using a matrix comprised of stool specimens negative for all assay targets preserved in Cary-Blair media processed into STM. Positive panel members were created by spiking 1.5X LoD (low positive) or 3X LoD (moderate positive) concentrations of the target analytes into the negative matrix.

The agreement with expected results was 100% for all panel members for *Yersinia*, *Vibrio*, STEC O157, and *Plesiomonas* (Table 9).

Table 9: Agreement of Panther Fusion GI Expanded Bacterial Assay Results with Expected Results

Description	Analyte	Agreement with Expected Results	
		N	% (95% CI)
Neg	Internal Control	89/89	100 (95.9-100)
Low Pos ^a	<i>Yersinia</i> ^c	90/90	100 (95.9-100)
	<i>Vibrio</i> ^c	90/90	100 (95.9-100)
	STEC O157 ^c	90/90	100 (95.9-100)
	<i>Plesiomonas</i> ^c	90/90	100 (95.9-100)
Mod Pos ^b	<i>Yersinia</i> ^{c d}	90/90	100 (95.9-100)
	<i>Vibrio</i> ^c	90/90	100 (95.9-100)
	STEC O157 ^c	90/90	100 (95.9-100)
	<i>Plesiomonas</i> ^c	90/90	100 (95.9-100)

CI = score confidence interval, Mod = moderate, N = sample size, Neg = negative, Pos = positive.

^a Low Pos = All targets are 1.5X LoD.

^b Mod Pos = All targets are 3X LoD.

^c *Yersinia enterocolitica*, *Vibrio parahaemolyticus*, STEC O157, and *Plesiomonas shigelloides* strains were used to build the positive panels.

^d One (1) false positive *Vibrio* result was obtained for a moderate positive *Yersinia* panel member.

Signal variability was measured as %CV of the Ct values. The total signal variability was ≤1.61% (SD ≤0.55) for all panel components (Table 10). For the sources of variation except the 'within run' factor, %CV values were ≤1.03% for all panel components. The signal variability was ≤1.01% (SD ≤0.33) for the Panther Fusion GI Expanded Bacterial Assay positive controls (Table 11).

Table 10: Signal Variability of the Panther Fusion GI Expanded Bacterial Assay by Target and Concentration

Description	Analyte	N	Mean Ct	Between Site		Between Operator/Run ^c		Between Day		Within Run		Total	
				SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)
Low Pos ^a	<i>Yersinia</i>	90	34.7	0.17	0.50	0.21	0.61	0.09	0.27	0.44	1.25	0.52	1.51
	<i>Vibrio</i>	90	33.7	0.16	0.49	0.08	0.25	0.00	0.00	0.26	0.77	0.32	0.95
	STEC O157	90	32.4	0.17	0.53	0.13	0.41	0.00	0.00	0.30	0.92	0.37	1.14
	<i>Plesiomonas</i>	90	33.9	0.16	0.47	0.06	0.17	0.00	0.00	0.32	0.94	0.36	1.06
Mod Pos ^b	<i>Yersinia</i>	90	33.8	0.35	1.03	0.19	0.58	0.07	0.21	0.37	1.08	0.55	1.61
	<i>Vibrio</i>	90	32.7	0.20	0.60	0.09	0.26	0.11	0.35	0.22	0.68	0.33	1.01
	STEC O157	90	31.4	0.24	0.75	0.08	0.27	0.07	0.21	0.26	0.81	0.36	1.16
	<i>Plesiomonas</i>	90	33.2	0.22	0.67	0.12	0.37	0.00	0.00	0.26	0.78	0.36	1.09

Ct = cycle threshold, CV = coefficient of variation, Mod = moderate, N = sample size, Pos = positive, SD = standard deviation.

Note: The analysis was performed using the SAS MIXED procedure, which applies a lower boundary of 0 to all variance components in the model by default. If a variance component is 0, SD, and %CV are displayed as 0.00

^a Low Pos = All targets are 1.5X LoD.

^b Mod Pos = All targets are 3X LoD.

^c Between Operator may be confounded with Between Run; therefore, Between Operator and Between Run estimates are combined in Between Operator/Run.

Table 11: Signal Variability of the Panther Fusion GI Expanded Bacterial Assay Positive Controls

Control	Analyte	N	Mean Ct	Between Site		Between Operator		Between Day		Within Day		Total	
				SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)	SD	CV (%)
Pos	<i>Yersinia</i>	30	32.7	0.22	0.66	0.00	0.00	0.00	0.00	0.25	0.75	0.33	1.01
	<i>Vibrio</i>	30	33.4	0.00	0.00	0.00	0.00	0.00	0.00	0.29	0.86	0.29	0.86
	STEC O157	30	31.5	0.11	0.35	0.00	0.00	0.07	0.23	0.26	0.83	0.29	0.93
	<i>Plesiomonas</i>	30	32.9	0.05	0.16	0.08	0.23	0.12	0.37	0.24	0.74	0.29	0.87

Ct = cycle threshold, CV = coefficient of variation, N = sample size, Pos = positive, SD = standard deviation.

Note: The analysis was performed using the SAS MIXED procedure, which applies a lower boundary of 0 to all variance components in the model by default. If a variance component is 0, SD and %CV are displayed as 0.00.

Clinical Performance

A multicenter study was conducted using remnant stool specimens in Cary-Blair preservative medium collected as part of routine patient care at 10 US clinics from pediatric or adult patients suspected of acute gastroenteritis. All specimens were tested with the Panther Fusion GI Expanded Bacterial Assay and with comparator assays: a PCR plus bidirectional sequencing (run in duplicate) for STEC O157 and an FDA-cleared Nucleic Acid Amplification Test (NAAT) for all other targets. An alternate FDA-cleared NAAT was used for discordant resolution testing, if applicable. Positive (PPA) and negative (NPA) percent agreement, with corresponding 2-sided 95% Score CIs, were calculated relative to comparator results, by target and by specimen category.

A total of 1,548 prospective specimens and 251 retrospective specimens were enrolled in the study; 94 specimens were excluded from the performance analyses (for example, duplicate individuals, invalid Panther Fusion GI Expanded Bacterial or comparator results for all targets). An additional 189 contrived specimens were assessed to supplement the prospective and retrospective data for all targets. Of the 1,919 specimens tested in valid Panther Fusion GI Expanded Bacterial Assay runs, 36 (1.9%) had initial invalid results. Upon retest, 25 of the 36 specimens yielded valid results, for a total of 11 (0.6%) specimens with final invalid results. The final data set consisted of 1,894 evaluable specimens (1,523 prospective specimens, 182 retrospective specimens, and 189 contrived specimens); not all were evaluable for all analytes. Demographic information for the 1,705 evaluable prospective and retrospective specimens is provided in Table 12.

Table 12: Summary of Subject Demographics

		Total N (%)	Prospective N (%)	Retrospective N (%)
Total Specimens		1,705	1,523	182
Sex	Female	888 (52.1)	793 (52.1)	95 (52.2)
	Male	817 (47.9)	730 (47.9)	87 (47.8)
Age Group	0 to 28 days	7 (0.4)	7 (0.5)	0 (0)
	29 days to <2 years	74 (4.3)	67 (4.4)	7 (3.8)
	2 to 5 years	55 (3.2)	50 (3.3)	5 (2.7)
	6 to 11 years	68 (4.0)	66 (4.3)	2 (1.1)
	12 to 17 years	73 (4.3)	71 (4.7)	2 (1.1)
	18 to 21 years	47 (2.8)	44 (2.9)	3 (1.6)
	22 to 64 years	825 (48.4)	724 (47.5)	101 (55.5)
	≥65 years	556 (32.6)	494 (32.4)	62 (34.1)

N = population size.

Performance characteristics for detection of *Yersinia*, *Vibrio*, STEC O157, and *Plesiomonas* are shown in Table 13 through Table 16.

Table 13: Clinical Performance - *Yersinia* spp.

Specimen Origin	N	TP	FP	TN	FN	Prevalence ^a (%)	PPA % (95% CI) ^b	NPA % (95% CI) ^b
Prospective (Fresh)	1,507	10	9 ^c	1,487	1 ^d	0.7	90.9 (62.3, 98.4)	99.4 (98.9, 99.7)
Retrospective (Frozen)	182	15	3 ^e	164	0	N/A ^f	100 (79.6, 100)	98.2 (94.9, 99.4)
Contrived (Frozen)	189	63	0	126	0	N/A ^f	100 (94.3, 100)	100 (97.0, 100)

CI = confidence interval, FN = false negative, FP = false positive, N = sample size, NPA = negative percent agreement, PPA = positive percent agreement, TN = true negative, TP = true positive.

^a Study prevalence reported based on comparator testing.

^b Score CI.

^c 6 of 9 discordant false positive prospective specimens were positive for *Yersinia* by the alternate NAAT.

^d The discordant false negative prospective specimen was negative for *Yersinia* by the alternate NAAT.

^e The 3 discordant false positive retrospective specimens were positive for *Yersinia* by the alternate NAAT.

^f Calculation of prevalence is not applicable.

Table 14: Clinical Performance - *Vibrio* spp.

Specimen Origin	N	TP	FP	TN	FN	Prevalence ^a (%)	PPA % (95% CI) ^b	NPA % (95% CI) ^b
Prospective (Fresh)	1,507	1	0	1,505	1 ^c	0.1	50.0 (9.5, 90.5)	100 (99.7, 100)
Retrospective (Frozen)	182	9	6 ^d	167	0	N/A ^f	100 (70.1, 100)	96.5 (92.6, 98.4)
Contrived (Frozen)	189	63	1 ^e	125	0	N/A ^f	100 (94.3, 100)	99.2 (95.6, 99.9)

CI = confidence interval, FN = false negative, FP = false positive, N = sample size, NPA = negative percent agreement, PPA = positive percent agreement, TN = true negative, TP = true positive.

^a Study prevalence reported based on comparator testing.

^b Score CI.

^c The discordant false negative prospective specimen was positive for *Vibrio* by the alternate NAAT.

^d All 6 discordant false positive retrospective specimens were positive for *Vibrio* by the alternate NAAT.

^e The discordant false positive contrived specimen was negative for *Vibrio* by the alternate NAAT.

^f Calculation of prevalence is not applicable.

Table 15: Clinical Performance - STEC O157

Specimen Origin	N	TP	FP	TN	FN	Prevalence ^a (%)	PPA % (95% CI) ^b	NPA % (95% CI) ^b
Prospective (Fresh)	1,522	1	2 ^c	1,519	0	0.1	100 (20.7, 100)	99.9 (99.5, 100)
Retrospective (Frozen)	182	3	1 ^d	178	0	N/A ^g	100 (43.9, 100)	99.4 (96.9, 99.9)
Contrived (Frozen)	189	62	1 ^e	125	1 ^f	N/A ^g	98.4 (91.5, 99.7)	99.2 (95.6, 99.9)

CI = confidence interval, FN = false negative, FP = false positive, N = sample size, NPA = negative percent agreement, PPA = positive percent agreement, TN = true negative, TP = true positive.

^a Study prevalence reported based on comparator testing.

^b Score CI.

^c 1 of 2 discordant false positive prospective specimens was negative for STEC O157 by the alternate NAAT. The other discordant was positive for O157 but negative for *stx1/stx2* by the alternate NAAT.

^d The discordant false positive retrospective specimen was positive for STEC O157 by the alternate NAAT.

^e The discordant false positive contrived specimen was negative for STEC O157 by the alternate NAAT.

^f The discordant false negative contrived specimen was not retested by the alternate NAAT.

^g Calculation of prevalence is not applicable.

Table 16: Clinical Performance - *Plesiomonas*

Specimen Origin	N	TP	FP	TN	FN	Prevalence ^a (%)	PPA % (95% CI) ^b	NPA % (95% CI) ^b
Prospective (Fresh)	1,507	1	1 ^c	1,505	0	0.1	100 (20.7, 100)	99.9 (99.6, 100)
Retrospective (Frozen)	182	8	1 ^d	173	0	N/A ^f	100 (67.6, 100)	99.4 (96.8, 99.9)
Contrived (Frozen)	189	62	0	126	1 ^e	N/A ^f	98.4 (91.5, 99.7)	100 (97.0, 100)

CI = confidence interval, FN = false negative, FP = false positive, N = sample size, NPA = negative percent agreement, PPA = positive percent agreement, TN = true negative, TP = true positive.

^a Study prevalence reported based on comparator testing.

^b Score CI.

^c The discordant false positive prospective specimen was positive for *Plesiomonas* by the alternate NAAT.

^d The discordant false positive retrospective specimen was positive for *Plesiomonas* by the alternate NAAT.

^e The discordant false negative contrived specimen was not retested by the alternate NAAT.

^f Calculation of prevalence is not applicable.

No coinfections were detected by the Panther Fusion GI Expanded Bacterial Assay or by the comparator methods in prospective and retrospective specimens.

Bibliography

1. WHO's first ever global estimates of foodborne diseases find children under 5 account for almost one third of deaths. Published December 3, 2015. Accessed May 27, 2025. <https://www.who.int/news/item/03-12-2015-who-s-first-ever-global-estimates-of-foodborne-diseases-find-children-under-5-account-for-almost-one-third-of-deaths>
2. Centers for Disease Control and Prevention. Published date unknown. Burden of foodborne illness: Overview. U.S. Department of Health & Human Services. Retrieved May 27, 2025, from https://archive.cdc.gov/www_cdc_gov/foodborneburden/estimates-overview.html
3. Riddle MS, DuPont HL, Connor BA. ACG Clinical Guideline: Diagnosis, Treatment, and Prevention of Acute Diarrheal Infections in Adults. *Am J Gastroenterol*. 2016;111(5):602–622. doi:10.1038/ajg.2016.141
4. Scallan E, Hoekstra RM, Angulo FJ, Tauxe RV, Widdowson MA, Roy SL, Jones JL, Griffin PM. Foodborne illness acquired in the United States—major pathogens. *Emerg Infect Dis*. 2011;17(1):7–15. doi:10.3201/eid1701.P11101
5. Ong KL, Gould LH, Chen DL, Jones TF, Scheftel J, Webb TH, Mody RK, Mahon BE. Changing epidemiology of *Yersinia enterocolitica* infections: markedly decreased rates in young black children, Foodborne Diseases Active Surveillance Network (FoodNet), 1996-2009. *Clin Infect Dis*. 2012 Jun;54 Suppl 5(0 5):S385-90. doi: 10.1093/cid/cis053.
6. Centers for Disease Control and Prevention. About Vibrio Infection. CDC. Updated May 14, 2024. Accessed June 2, 2025. <https://www.cdc.gov/vibrio/about/index.html>
7. Armed Forces Health Surveillance Division. *Escherichia coli*, Shiga Toxin-Producing (STEC) Reference Sheet. U.S. Department of Defense; 2022. Accessed May 30, 2025. <https://ph.health.mil/cdt/cphe-cdt-e-coli-shiga-toxin-producing-ref.pdf>
8. Morris JG Jr, Horneman A. *Plesiomonas shigelloides* infections. *UpToDate*. Calderwood SB, Baron EL, eds. Updated December 13, 2023. Accessed May 30, 2025. <https://www.uptodate.com/contents/plesiomonas-shigelloides-infections/print>
9. Clinical and Laboratory Standards Institute (CLSI). Protection of Laboratory Workers from Occupationally Acquired Infections, 3rd ed. CLSI guideline M29. Clinical and Laboratory Standards Institute; 2014. Accessed May 30, 2025. <https://clsi.org/standards/products/microbiology/documents/m29/>

Contact Information



Hologic, Inc.
10210 Genetic Center Drive
San Diego, CA 92121 USA

For country-specific Technical Support and Customer Service, email address and telephone number, visit www.hologic.com/support.

Hologic, Aptima, Panther, Panther Fusion, and associated logos are trademarks and/or registered trademarks of Hologic, Inc. and/or its subsidiaries in the United States and/or other countries.

All other trademarks that may appear in this package insert are the property of their respective owners.

This product may be covered by one or more U.S. patents identified at www.hologic.com/patents.

©2025 Hologic, Inc. All rights reserved.

AW-33558-001 Rev. 001
2025-10