

Updated X-ray Generator Specifications

This addendum modifies the X-ray Generator specifications for your Affirm® Prone biopsy system *User Guide* (MAN-10818 Revision 001) and *Service Manual* (MAN-10809 Revision 001). Specifically, the *Rating* and *mAs Range* specifications were clarified.

Place this addendum with your product manuals for future reference.

X-ray Generator

<i>Type</i>	<i>Constant potential, three-phase, high frequency inverter</i>
<i>Rating</i>	<i>7.0 kW maximum (200 mA at 35 kVp)</i>
<i>Electrical Power Capacity</i>	<i>9.0 kW maximum</i>
<i>kV Range</i>	<i>20-49 kVp in 1 kVp increments</i>
<i>kV Accuracy</i>	<i>±5%</i>
<i>mA Range</i>	<i>10-200 mA</i>
<i>mAs Accuracy</i>	<i>±(10% + 0.2 mAs)</i>
<i>mAs Range</i>	<i>(2D) 4.0 – 500 mAs</i> <i>(3D) 6.0 – 300 mAs</i>