













## Breast and Skeletal Health Products and Labels






### Glossary of Symbols

This is a glossary of symbols on Hologic Breast and Skeletal Health products and labels. Use this glossary as a supplement to your product manuals.

Symbol	Description	Standard
	Prescription use only	FDA 21 CFR 801.109
	Not safe for magnetic resonance imaging	ASTM F2503 Reference no. Table 2, Symbol 7.3.3;7.4.9.1; Fig. 9
	Conditional use for magnetic resonance imaging	ASTM F2503 Reference no. Table 2; 7.4.6.1; Fig 6,7
	Caution—radiation	Hologic
	Emergency stop	Hologic
	Wheelchair accessible	Hologic
	Device driver and remote-control connection	Hologic
	Device driver cord management plate	Hologic
	Footswitch connection	Hologic
	Footswitch cord management plate	Hologic
	Saline tubing management plate	Hologic
	Wi-Fi	Hologic

## Breast and Skeletal Health Products and Labels

### Glossary of Symbols

Symbol	Description	Standard
	Danger: Pinch Point	Hologic
	Bluetooth	Hologic
	Battery	Hologic
<p><b>YYYY-MM-DD</b></p>	Date format: YYYY represents the year MM represents the month DD represents the day	Hologic
	Circuit breaker	Hologic
	Place equipment outside of patient environment.	Hologic
<p><b>Translations in Box</b></p>	Translations in Box	Hologic
<p><b>Temp Logger Activated</b></p>	Temperature Logger activated	Hologic
<p><b>Temp Logger SN</b></p>	Temperature Logger Serial Number	Hologic
<p><b>P/N</b></p>	Part Number	Hologic
<p><b>REV</b></p>	Revision	Hologic
<p><b>Model</b></p>	Model Number	Hologic
<p><b>QTY</b></p>	Quantity	Hologic
<p><b>SFW SHIPPED</b></p>	Software version shipped	Hologic
















## Breast and Skeletal Health Products and Labels

### Glossary of Symbols

Symbol	Description	Standard
<b>SFW Version</b>	Software version number	Hologic
<b>APPLICATION INSTALLATION</b>	Application Installation	Hologic
<b>Verified By</b>	Verified By	Hologic
<b>CASE</b>	Case (box or package)	Hologic
<b>VC</b>	Vendor Code	Hologic
<b>Made in USA</b>	Made in United States of America	Hologic
<b>Made in Costa Rica</b>	Made in Costa Rica	Hologic
<b>Assembled in Costa Rica</b>	Assembled in Costa Rica	Hologic
<b>Made in China</b>	Made in China	Hologic
<b>Made in Mexico</b>	Made in Mexico	Hologic
<b>Made in Thailand</b>	Made in Thailand	Hologic
<b>Patents</b>	Patents	Hologic
<b>USA REP</b>	USA Representative	Hologic
<b>LATERAL ARM USE ONLY</b>	Lateral Arm Use Only	Hologic
<b>Q.C. Use Only</b>	Quality Control Use only	Hologic
<b>UNLOAD THIS SIDE</b>	Unload this side	Hologic
<b>RAMP THIS SIDE</b>	Ramp this side	Hologic











## Breast and Skeletal Health Products and Labels

### Glossary of Symbols

Symbol	Description	Standard
	Unlocked	IEC 60417 Reference no. 5569
	"ON" (power)	IEC 60417, Reference 5007
	"OFF" (power)	IEC 60417, Reference 5008
	Fuse	IEC 60417, Reference 5016
	Protective earth (ground)	IEC 60417, Reference 5019
	Potential equalization terminal	IEC 60417, Reference 5021
	Alternating Current	IEC 60417, Reference 5032
	Dangerous voltage	IEC 60417, Reference 5036
	This system transmits radio frequency (RF) energy (non-ionizing electromagnetic radiation)	IEC 60417, Reference 5140
	Type BF applied part	IEC 60417, Reference 5333
	Type B applied part	IEC 60417, Reference 5840
	Ethernet	IEC 60417, Reference 5988
	Warning electricity	IEC 60417, Reference 6042
	USB	IEC 60417-1
	Locked	IEC 60417-1 ISO 7000-5569











## Breast and Skeletal Health Products and Labels

### Glossary of Symbols

Symbol	Description	Standard
	"On" and "Off" (power) for the computer and display.	IEC 60417-5010
	Earth(ground)	IEC 60417-5017
	Electrostatic sensitive device	IEC 60417-5134
	"ON" for part of the equipment	IEC 60417-5264
	Computer On switch	IEC 60417-5264
	Computer Standby switch	IEC 60417-5266
	X-ray source assembly	IEC 60417-5338
	X-ray source emitting	IEC 60417-5339
	Radiation filter	IEC 60417-5381
	Do not immerse in any liquid	IEC 60417-5995
<b>IPX8</b>	The equipment of accessory is suitable for continuous immersion in water (up to 1m of submersion for 1 hour).	IEC 60529














## Breast and Skeletal Health Products and Labels

### Glossary of Symbols

Symbol	Description	Standard
 <b>IPX4</b>	Splash -proof	IEC 60529
<b>IPX6</b>	Water projected in powerful jets against the equipment or accessory from any direction shall have no harmful effects.	IEC 60529
<b>IP54</b>	Protected against dust limited ingress, no harmful deposits and Protected against water splashed from all directions, limited ingress permitted	IEC 60529
<b>IP22</b>	Protected against solid foreign objects of 12,5 mm Ø and greater And Protection against vertically falling water drops when ENCLOSURE tilted up to 15°	IEC 60529
	Medical device	ISO 15223 Draft
	Manufacturer	ISO 15223-1:2016, Reference 5.1.1
	Authorized Representative in the European Community	ISO 15223-1:2016, Reference 5.1.2
	Date of manufacture	ISO 15223-1:2016, Reference 5.1.3
	Use-by Date	ISO 15223-1:2016, Reference 5.1.4
	Batch code	ISO 15223-1:2016, Reference 5.1.5
	Catalog number	ISO 15223-1:2016, Reference 5.1.6
	Serial number	ISO 15223-1:2016, Reference 5.1.7
	Sterilized using ethylene oxide	ISO 15223-1:2016, Reference 5.2.3














## Breast and Skeletal Health Products and Labels

### Glossary of Symbols

Symbol	Description	Standard
	Sterilized using irradiation	ISO 15223-1:2016, Reference 5.2.4
	Do not re sterilize	ISO 15223-1:2016, Reference 5.2.6
	Non-sterile	ISO 15223-1:2016, Reference 5.2.7
	Do not use if package is damaged	ISO 15223-1:2016, Reference 5.2.8
	Fragile; handle with care	ISO 15223-1:2016, Reference 5.3.1
	Keep dry	ISO 15223-1:2016, Reference 5.3.4
	Lower limit of temperature	ISO 15223-1:2016, Reference 5.3.5
	Upper limit of temperature	ISO 15223-1:2016, Reference 5.3.6
	Temperature limit	ISO 15223-1:2016, Reference 5.3.7
	Humidity limitation	ISO 15223-1:2016, Reference 5.3.8
	Atmospheric pressure limitation	ISO 15223-1:2016, Reference 5.3.9
	Do not re-use	ISO 15223-1:2016, Reference 5.4.2
 <a href="http://www.hologic.com/package-inserts">www.hologic.com/package-inserts</a>	Consult Electronic Instructions for use	ISO 15223-1:2016, Reference 5.4.3

## Breast and Skeletal Health Products and Labels





### Glossary of Symbols

Symbol	Description	Standard
	Caution	ISO 15223-1:2016, Reference 5.4.4
	Combined weight of the equipment and its safe working load	ISO 60417, Reference 1321A
	This way up	ISO 7000:2014, Reference 0623
	Do not stack	ISO 7000:2014, Reference 2402
	Stacking limit by number	ISO 7000:2014, Reference 2403
	Pinch point	ISO -7010
	Warning	ISO 7010, Reference W001
	CE Mark with Notified Body reference number	MDR Regulation (EU) 2017/745
	Contents	N/A
	Does not contain natural rubber latex	N/A
	Universal Serial Bus connection	N/A
	Do not cut	N/A
	CSA Listed device	Standards Council of Canada (SCC) and the U.S. Occupational Safety and Health Administration (OSHA)



## Breast and Skeletal Health Products and Labels

### Glossary of Symbols

Symbol	Description	Standard
	<p>Medical - General medical equipment as to electrical shock, fire, and mechanical hazards only in accordance with ANSI/AAMI ES 60601-1:2005 (AI:2012), CAN/CSA C22.2 No. 60601-1 (2014), and IEC 60601-1:2012.</p>	<p>UL Classification</p>
	<p>INMETRO &amp; ULBR Mark</p>	<p>UL-BR INMETRO certification (Brazil)</p>
	<p>Recyclable symbol</p>	<p>Unicode, Reference U+267C</p>
	<p>Discard electrical and electronic equipment separately from standard waste. Send decommissioned material to Hologic or contact your service representative.</p>	<p>WEEE Directive 2002/96/EC</p>