

AdV/hMPV/RV Assay (Panther Fusion® System)

For *in vitro* diagnostic use.

Rx Only.

CONTENTS

General Information 2
 Intended Use 2
 Summary and Explanation of the Test. 2
 Principles of the Procedure 3
 Warnings and Precautions. 4
 Reagent Storage and Handling Requirements 6
 Specimen Collection and Storage 7
 Specimen Transport 8
Panther Fusion System 9
 Reagents and Materials Provided for Panther Fusion AdV/hMPV/RV Assay 9
 Materials Required and Available Separately 10
 Panther Fusion System Test Procedure 11
 Procedural Notes 12
Quality Control 12
Interpretation of Results. 13
Limitations 14
Panther Fusion System Assay Performance 15
 Reproducibility 15
 Clinical Performance 16
 Analytical Sensitivity 18
 Reactivity 19
 Analytical Specificity 21
 Competitive Interference 23
 Interference 23
 Carry-Over/Contamination. 24
 Assay Precision 25
Bibliography 27

General Information

Intended Use

The Panther Fusion® AdV/hMPV/RV assay is a multiplex real-time PCR (RT-PCR) *in vitro* diagnostic test for the rapid and qualitative detection and differentiation of Adenovirus (AdV), human Metapneumovirus (hMPV), and Rhinovirus (RV). Nucleic acids are isolated and purified from nasopharyngeal (NP) swab specimens obtained from individuals exhibiting signs and symptoms of a respiratory tract infection.

This assay is intended to aid in the differential diagnosis of Adenovirus, human Metapneumovirus, and Rhinovirus infections in humans. Negative results do not preclude Adenovirus, human Metapneumovirus, and Rhinovirus infections and should not be used as the sole basis for treatment or other management decisions. This assay is designed for use on the Panther Fusion system.

Summary and Explanation of the Test

Respiratory viruses are responsible for a wide range of acute respiratory tract infections including the common cold, influenza, and croup and represent the most common cause of acute illness in the United States. Disease severity can be especially high in the young, the immunocompromised, and elderly patients. Accurate and timely diagnosis of the cause of respiratory tract infections has many benefits. They include improved treatment of the patient by ensuring appropriate antiviral treatment (e.g. oseltamivir for influenza), decreasing the overall cost of care, reducing selection for antimicrobial resistant organisms due to excessive and inappropriate use of antibiotics,¹ assisting infection control personnel in providing appropriate measures to minimize nosocomial spread, and providing valued information to public health authorities regarding which viruses are circulating in the community.²

Adenoviruses are members of the *Adenoviridae* family which are medium-sized (90-100 nm), non-enveloped icosahedral viruses with double-stranded DNA.³ At this time, in humans there are over 50 Adenovirus types in seven species (A to G).⁴ Adenoviruses most commonly cause respiratory illness which can range from the common cold to pneumonia, croup, and bronchitis.³ Depending on the type, Adenoviruses can cause other illnesses such as gastroenteritis, conjunctivitis, cystitis, and, less commonly, neurological disease.³ Infants and people with weakened immune systems are at high risk for developing severe illness caused by Adenovirus infection.³ Adenovirus circulates year-round and outbreaks are more common in late winter, spring, and early summer but can occur throughout the year.⁵

Since the discovery of hMPV in 2001, the virus has been identified worldwide. hMPV is a common respiratory pathogen, particularly in infants and young children. The virus is associated with both upper and lower respiratory tract infections and may be a trigger for asthma.⁶ Symptoms commonly associated with hMPV include cough, fever, nasal congestion, and shortness of breath. Clinical symptoms of hMPV infection may progress to bronchiolitis or pneumonia and are similar to other viruses that cause upper and lower respiratory infections. The incubation period is estimated to be 3 to 6 days, and the median duration of illness can vary depending upon severity but is similar to other respiratory infections caused by viruses.⁷ The peak of incidence of hMPV is mainly in the spring in temperate latitudes.⁸

Rhinoviruses, members of the family Picornaviridae, are the causative pathogens in more than half of viral respiratory infections, and they are associated with acute exacerbations of

respiratory disease, including asthma, sinusitis, otitis media, and COPD.⁹ A number of studies have confirmed rhinoviruses as being the most common cause of “the common cold” and affect all age groups.⁸ Symptoms usually include sore throat, runny nose, coughing, sneezing, watery eyes, headaches and body aches. Most people recover within about 7-10 days.⁸ Rhinoviruses circulate year round and tend to peak in the spring and fall.⁸

Principles of the Procedure

The Panther Fusion AdV/hMPV/RV assay involves the following steps: sample lysis, nucleic acid capture and elution transfer, and multiplex RT-PCR when analytes are simultaneously amplified, detected and differentiated. Nucleic acid capture and elution takes place in a single tube on the Panther Fusion system. The eluate is transferred to the Panther Fusion system reaction tube containing the assay reagents. Multiplex RT-PCR is then performed for the eluted nucleic acid on the Panther Fusion system.

Nucleic acid capture and elution: Prior to processing and testing on the Panther Fusion system, specimens are transferred to a Specimen Lysis Tube containing specimen transport media (STM) that lyses the viral particles, releases target nucleic acid and protects it from degradation during storage.

The Internal Control-S (IC-S) is added to each test specimen and controls via the working Panther Fusion Capture Reagent-S (wFCR-S). The IC-S in the reagent monitors specimen processing, amplification and detection.

Capture oligonucleotides hybridize to nucleic acid in the test specimen. Hybridized nucleic acid is then separated from the specimen in a magnetic field.

Wash steps remove extraneous components from the reaction tube. The elution step elutes purified nucleic acid. During the nucleic acid capture and elution step, total nucleic acid is isolated from specimens.

Elution transfer and RT-PCR: During the elution transfer step, eluted nucleic acid is transferred to a Panther Fusion reaction tube already containing oil and reconstituted mastermix.

For RV, hMPV, and internal control targets, amplification occurs via RT-PCR. A reverse transcriptase step generates DNA copies of the target sequence. For AdV, target amplification occurs via PCR. For all targets, specific forward and reverse primers and probes amplify targets while simultaneously detecting and discriminating multiple target types via multiplex PCR.

The Panther Fusion system compares the fluorescence signal to a predetermined cut-off to produce a qualitative result for the presence or absence of the analyte.




The analytes and the channel used for their detection on the Panther Fusion system is summarized in the table below.

| Analyte | Gene Targeted | Instrument Channel |
|-----------------------|----------------|--------------------|
| Adenovirus | Hexon | HEX |
| human Metapneumovirus | Nucleocapsid | ROX |
| Rhinovirus | 5' UTR | FAM |
| Internal Control | Not applicable | RED677 |

Warnings and Precautions

- A. For *in vitro* diagnostic use.
- B. Carefully read this entire package insert and the *Panther Fusion System Operator's Manual*.
- C. The Panther Fusion Enhancer Reagent-S (FER-S) is corrosive, harmful if swallowed and causes severe skin burns and eye damage.
- D. Only personnel adequately trained on the use of this assay and in handling potentially infectious materials should perform these procedures. If a spill occurs, immediately disinfect using appropriate site procedures.
- E. Specimens may be infectious. Use Universal Precautions when performing this assay. Proper handling and disposal methods should be established by the laboratory director. Only personnel adequately trained in handling infectious materials should be permitted to perform this diagnostic procedure.¹⁰
- F. Use only supplied or specified disposable laboratory ware.
- G. Wear disposable, powderless gloves, protective eye wear, and laboratory coats when handling specimens and reagents. Wash hands thoroughly after handling specimens and reagents.
- H. Dispose of all material that has come into contact with specimens and reagents in accordance with applicable national, international, and regional regulations.
- I. Expiration dates listed on the Panther Fusion Specimen Lysis Tubes pertain to the transfer of sample into the tube and not to testing of the sample. Specimens collected/transferred any time prior to these expiration dates are valid for testing provided they are transported and stored in accordance with the appropriate package insert, even if these expiration dates have passed.
- J. Maintain proper storage conditions during specimen shipping to ensure the integrity of the specimen. Specimen stability under shipping conditions other than those recommended has not been evaluated.
- K. Avoid cross-contamination during the specimen handling steps. Specimens can contain extremely high levels of virus or other organisms. Ensure that specimen containers do not come in contact with one another, and discard used materials without passing them over any open containers. Change gloves if they come in contact with specimens.
- L. Do not use the reagents and controls after the expiration date.
- M. Store assay components at the recommended storage condition. See *Reagent Storage and Handling Requirements* (page 6), and *Panther Fusion System Test Procedure* (page 11) for more information.
- N. Do not combine any assay reagents or fluids. Do not top off reagents or fluids; the Panther Fusion system verifies reagent levels.
- O. Avoid microbial and ribonuclease contamination of reagents.

- P. Quality control requirements must be performed in conformance with local, state, and/or federal regulations or accreditation requirements and your laboratory's standard quality control procedures.
- Q. Do not use the assay cartridge if the storage pouch has lost its seal or if the assay cartridge foil is not intact. Contact Hologic if either occurs.
- R. Do not use the fluid packs if the foil seal is leaking. Contact Hologic if this occurs.
- S. Handle the assay cartridges with care. Do not drop or invert assay cartridges. Avoid prolonged exposure to ambient light.

| | |
|--|--|
|  | Panther Fusion Oil <i>Polydimethylsiloxane 95-100%</i> |
| | Warning H315 - Causes skin irritation H319 - Causes serious eye irritation P264 - Wash face, hands and any exposed skin thoroughly after handling P280 - Wear protective gloves/protective clothing/eye protection/face protection P305 + P351 + P338 - IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing P337 + P313 - If eye irritation persists: Get medical advice/attention P302 + P352 - IF ON SKIN: Wash with plenty of soap and water P332 + P313 - If skin irritation occurs: Get medical advice/attention P362 - Take off contaminated clothing and wash before reuse |
|   | Panther Fusion Enhancer Reagent-S <i>Lithium Hydroxide Monohydrate 5-10%</i> |
| | Danger H302 - Harmful if swallowed H314 - Causes severe skin burns and eye damage P260 - Do not breathe dust/fume/gas/mist/vapours/spray P264 - Wash face, hands and any exposed skin thoroughly after handling P270 - Do not eat, drink or smoke when using this product P280 - Wear protective gloves/protective clothing/eye protection/face protection P301 + P312 - IF SWALLOWED: Call a POISON CENTER or doctor/physician if you feel unwell P301 + P330 + P331 - IF SWALLOWED: rinse mouth. Do NOT induce vomiting P303 + P361 + P353 - IF ON SKIN (or hair): Remove/Take off immediately all contaminated clothing. Rinse skin with water/shower P304 + P340 - IF INHALED: Remove victim to fresh air and keep at rest in a position comfortable for breathing P305 + P351 + P338 - IF IN EYES: Rinse cautiously with water for several minutes. Remove contact lenses, if present and easy to do. Continue rinsing P310 - Immediately call a POISON CENTER or doctor/physician P330 - Rinse mouth P363 - Wash contaminated clothing before reuse |

Note: For information on any hazard and precautionary statements that may be associated with reagents, refer to the Safety Data Sheet Library at www.hologic.com/sds.

Reagent Storage and Handling Requirements

A. The following table provides storage and handling requirements for this assay.

| Reagent | Unopened Storage | On Board/ Open Stability ¹ | Opened Storage |
|---|------------------|--|-------------------------------|
| Panther Fusion AdV/hMPV/RV Assay Cartridge | 2°C to 8°C | 60 days | 2°C to 8°C ² |
| Panther Fusion Capture Reagent-S (FCR-S) | 15°C to 30°C | 30 days | 15°C to 30°C |
| Panther Fusion Enhancer Reagent-S (FER-S) | 15°C to 30°C | 30 days | 15°C to 30°C |
| Panther Fusion Internal Control-S (IC-S) | 2°C to 8°C | (In wFCR-S) | Not applicable |
| Panther Fusion Elution Buffer | 15°C to 30°C | 60 days | 15°C to 30°C |
| Panther Fusion Oil | 15°C to 30°C | 60 days | 15°C to 30°C |
| Panther Fusion Reconstitution Buffer I | 15°C to 30°C | 60 days | 15°C to 30°C |
| Panther Fusion AdV/hMPV/RV Positive Control | 2°C to 8°C | Single use vial | Not applicable- single use |
| Panther Fusion Negative Control | 2°C to 8°C | Single use vial | Not applicable- single use |

When reagents are removed from the Panther Fusion system, return them immediately to their appropriate storage temperatures.

¹ On board stability starts at the time the reagent is placed on the Panther Fusion system for the Panther Fusion AdV/hMPV/RV assay cartridge, FCR-S, FER-S and IC-S. The on board stability for the Panther Fusion Reconstitution Buffer I, Panther Fusion Elution Buffer and Panther Fusion Oil starts when the reagent pack is first used.

² If removed from the Panther Fusion System, store the assay cartridge in an air-tight container with desiccant at the recommended storage temperature.

- B. Working Panther Fusion Capture Reagent-S and Panther Fusion Enhancer Reagent-S are stable for 60 days when capped and stored at 15°C to 30°C. Do not refrigerate.
- C. Discard any unused reagents that have surpassed their on board stability.
- D. Controls are stable until the date indicated on the vials.
- E. Avoid cross-contamination during reagent handling and storage.
- F. **Do not freeze reagents.**

Specimen Collection and Storage

Specimens - Clinical material collected from patient placed in an appropriate transport system. For the Panther Fusion AdV/hMPV/RV assay this includes NP swab specimens in viral transport medium (VTM).

Samples - Represents a more generic term to describe any material for testing on the Panther Fusion System including specimens, specimens transferred into a Panther Fusion Specimen Lysis Tube and controls.

Note: Handle all specimens as if they contain potentially infectious agents. Use Universal Precautions.

Note: Take care to avoid cross-contamination during specimen handling steps. For example, discard used material without passing over open tubes.

A. Specimen types include NP swab specimens.

Collect NP swab specimens according to standard technique using a polyester-, rayon-, or nylon-tipped swab. Immediately place the swab specimen into 3mL of VTM.

The following types of VTM were verified for use.

- Remel MicroTest M4, M4RT, M5 or M6 formulations
- Copan Universal Transport Medium
- BD Universal Viral Transport Medium

B. Specimen processing

1. Prior to testing on the Panther Fusion system, transfer specimen* to a Panther Fusion Specimen Lysis Tube.

- Transfer 500 µL of the NP swab specimens to a Panther Fusion Specimen Lysis Tube.

***Note:** When testing frozen specimen, allow specimen to reach room temperature prior to processing.

2. Storing specimens before testing

a. After collection, specimens can be stored at 2°C to 8°C up to 96 hours before transferred to the Panther Fusion Specimen Lysis Tube. Remaining specimen volumes can be stored at ≤-70°C.

b. Specimen in the Panther Fusion Specimen Lysis Tube may be stored under one of the following conditions:

- 15°C to 30°C up to 6 days or
- 2°C to 8°C up to 3 months.

Note: It is recommended that specimens transferred to the Panther Fusion Specimen Lysis Tube are stored capped and upright in a rack.

C. Specimen on board the Panther Fusion system may be archived for additional testing at a later time.

D. Storing samples after testing

1. Samples that have been assayed should be stored upright in the rack under one of the following conditions:
 - 15°C to 30°C up to 6 days or
 - 2°C to 8°C up to 3 months.
2. The samples should be covered with a new, clean plastic film or foil barrier.
3. If assayed samples need to be frozen or shipped, remove the penetrable cap and place a new non-penetrable cap on the specimen tubes. If samples need to be shipped for testing at another facility, recommended temperatures must be maintained. Prior to uncapping previously tested and recapped samples, specimen transport tubes must be centrifuged for 5 minutes at 420 Relative Centrifugal Force (RCF) to bring all of the liquid down to the bottom of the tube. Avoid splashing and cross-contamination.

Specimen Transport

Maintain specimen storage conditions as described in the *Specimen Collection and Storage section* on page 7.

Note: *Specimens must be shipped in accordance with applicable national, international, and regional transportation regulations.*

Panther Fusion System

The Panther Fusion System is an integrated nucleic acid testing system that fully automates all steps necessary to perform various Panther Fusion assays from sample processing through amplification, detection, and data reduction.

Reagents and Materials Provided for Panther Fusion AdV/hMPV/RV Assay

Assay Packaging

| Components ¹ | Part No. | Storage |
|---|-----------|--------------|
| Panther Fusion AdV/hMPV/RV Assay Cartridges 96 Tests Panther Fusion AdV/hMPV/RV assay cartridge, 12 tests, 8 per box | PRD-04330 | 2°C to 8°C |
| Panther Fusion Internal Control-S 960 Tests Panther Fusion Internal Control-S tube, 4 per box | PRD-04332 | 2°C to 8°C |
| Panther Fusion AdV/hMPV/RV Assay Controls Panther Fusion AdV/hMPV/RV Positive Control tube, 5 per box Panther Fusion Negative Control tube, 5 per box | PRD-04338 | 2°C to 8°C |
| Panther Fusion Extraction Reagent-S 960 Tests Panther Fusion Capture Reagent-S bottle, 240 tests, 4 per box Panther Fusion Enhancer Reagent-S bottle, 240 tests, 4 per box | PRD-04331 | 15°C to 30°C |
| Panther Fusion Elution Buffer 2400 Tests Panther Fusion Elution Buffer pack, 1200 tests, 2 per box | PRD-04334 | 15°C to 30°C |
| Panther Fusion Reconstitution Buffer I 1920 Tests Panther Fusion Reconstitution Buffer I pack, 960 tests, 2 per box | PRD-04333 | 15°C to 30°C |
| Panther Fusion Oil 1920 Tests Panther Fusion Oil pack, 960 tests, 2 per box | PRD-04335 | 15°C to 30°C |

¹ Components can also be ordered in the following bundles:

Panther Fusion Universal Fluids Kit, PRD-04430, contains 1 each Panther Fusion Oil and Panther Fusion Elution buffer.

Panther Fusion Assay Fluids I-S, PRD-04431, contains 2 Panther Fusion Extraction Reagents-S, 2 Panther Fusion Internal Control-S, and 1 Panther Fusion Reconstitution Buffer I.

Individually Packaged Items

| Items | Part No. |
|--|-----------|
| Panther Fusion Specimen Lysis Tubes, 100 per bag | PRD-04339 |

Materials Required and Available Separately

Note: Materials available from Hologic have catalog numbers listed, unless otherwise specified.

| Material | Cat. No. |
|---|---------------------------|
| Panther System | 303095 |
| Panther Fusion Module | ASY-09600 |
| Aptima Assay Fluids Kit (Aptima Wash Solution, Aptima Buffer for Deactivation Fluid, and Aptima Oil Reagent) | 303014 (1000 tests) |
| Multi-tube units (MTUs) | 104772-02 |
| Panther Waste Bag Kit | 902731 |
| Panther Waste Bin Cover | 504405 |
| Or Panther System Run Kit for Real Time Assays contains MTUs, waste bags, waste bin covers, and assay fluids | PRD-03455 (5000 tests) |
| Or Panther System Run Kit (when running TMA assays in parallel with real time-TMA assays) contains MTUs, waste bags, waste bin covers, auto detect*, and assay fluids | 303096 (5000 tests) |
| Panther Fusion Tube Trays, 1008 tests, 18 trays per box | PRD-04000 |
| Liquid Handling (LiHa) Disposable Tips, 1000 µL | 10612513 (Tecan) |
| Aptima penetrable caps (optional) | 105668 |
| Replacement non-penetrable caps (optional) | 103036A |
| Replacement extraction reagent bottle caps | CL0040 |
| P1000 pipettor and tips with hydrophobic plugs | - |
| Bleach, 5% to 7% (0.7 M to 1.0 M) sodium hypochlorite solution <i>Note: Mix one part bleach with one part deionized water to make diluted working bleach solution 2.5% to 3.5% (0.35 M to 0.5 M) sodium hypochlorite solution.</i> | - |
| Disposable powderless gloves | - |

*Needed only for Panther Aptima TMA assays.

Panther Fusion System Test Procedure

Note: Refer to the Panther Fusion System Operator's Manual for additional procedural information.

A. Work Area Preparation

1. Wipe down work surfaces with 2.5% to 3.5% (0.35 M to 0.5 M) sodium hypochlorite solution. Allow the sodium hypochlorite solution to contact surfaces for at least 1 minute and follow with a deionized (DI) water rinse. Do not allow the sodium hypochlorite solution to dry. Cover the bench surface with clean, plastic-backed absorbent laboratory bench covers.
2. Clean a separate work surface where samples will be prepared using the procedure described in step A.1.

B. Reagent Preparation

1. Remove the bottles of IC-S, FCR-S and FER-S from storage.
2. Open the bottles of IC-S, FCR-S and FER-S, and discard the caps. Open the TCR door on the upper bay of the Panther Fusion system.
3. Place the IC-S, FCR-S and FER-S bottles in the appropriate positions on the TCR carousel.
4. Close the TCR door.

Note: The Panther Fusion system adds the IC-S to the FCR-S. After the IC-S is added to the FCR-S, it is referred to as wFCR-S (working FCR-S). If the FCR-S and FER-S are removed from the system, use new caps and immediately store according to the proper storage conditions.

C. Specimen Handling

Note: Prepare specimens per the Specimen Processing instructions in the Specimen Collection and Storage section before loading specimens onto the Panther Fusion system.

1. **Do not vortex samples.**
2. Inspect sample tubes before loading into the rack. If a sample tube contains bubbles or has a lower volume than is typically observed, gently tap the bottom of the tube to bring contents to the bottom.

Note: To avoid a processing error, ensure adequate specimen volume is added to the Panther Fusion Specimen Lysis Tube. When 500 μ L of NP swab specimen is added to the Panther Fusion Specimen Lysis Tube, there is sufficient volume to perform 3 nucleic acid extractions.

D. System Preparation

For instructions on setting up the Panther Fusion system including loading samples, assay cartridges and universal fluids, refer to the Panther Fusion System Operator's Manual.

Procedural Notes

A. Controls

1. The Panther Fusion AdV/hMPV/RV Positive Control and Panther Fusion Negative Control can be loaded in any rack position, in any Sample Bay lane on the Panther Fusion system.
2. Once the control tubes are pipetted and are processed for the Panther Fusion AdV/hMPV/RV assay, they are active for up to 30 days (control frequency configured by an administrator) unless control results are invalid or a new assay cartridge lot is loaded.
3. Each control tube can be tested once.
4. Patient specimen pipetting begins when one of the following two conditions is met:
 - a. Valid results for the controls are registered on the system.
 - b. A pair of controls is currently in process on the system.

Quality Control

A run or specimen result may be invalidated by the Panther Fusion System if problems occur while performing the assay. Specimens with invalid results must be retested.

Negative and Positive Controls

To generate valid results, a set of assay controls must be tested. One replicate of the negative assay control and positive assay control must be tested each time a new lot of assay cartridges is loaded on the Panther Fusion system or when the current set of valid controls for an active cartridge lot have expired.

The Panther Fusion system is configured to require assay controls run at an administrator-specified interval of up to 30 days. Software on the Panther Fusion system alerts the operator when assay controls are required and does not start new tests until the assay controls are loaded and have started processing.

During processing, criteria for acceptance of the assay controls are automatically verified by the Panther Fusion system. To generate valid results, the assay controls must pass a series of validity checks performed by the Panther Fusion system.

If the assay controls pass all validity checks, they are considered valid for the administrator-specified time interval. When the time interval has passed, the assay controls are expired by the Panther Fusion system and requires a new set of assay controls be tested prior to starting any new samples.

If any one of the assay controls fails the validity checks, the Panther Fusion system automatically invalidates the affected samples and requires a new set of assay controls be tested prior to starting any new samples.

Internal Control

An internal control is added to each sample during the extraction process. During processing, the internal control acceptance criteria is automatically verified by the Panther Fusion system software. Detection of the internal control is not required for samples that are positive for AdV, hMPV and/or RV. The internal control must be detected in all samples that are negative for AdV, hMPV and RV targets; samples that fail to meet that criteria will be reported as Invalid. Each sample with an Invalid result must be retested.

The Panther Fusion system is designed to accurately verify processes when procedures are performed following the instructions provided in this package insert and the *Panther Fusion System Operator's Manual*.

Interpretation of Results

The Panther Fusion system automatically determines the test results for samples and controls. Results for AdV, hMPV and RV detection are reported separately. A test result may be negative, positive, or invalid.

Table 1 shows the possible results reported in a valid run with result interpretations.

Table 1: Result Interpretation

| AdV Result | hMPV Result | RV Result | IC Result | Interpretation |
|------------|-------------|-----------|-----------|---|
| Neg | Neg | Neg | Valid | AdV, hMPV and RV not detected. |
| POS | Neg | Neg | Valid | AdV detected. hMPV and RV not detected. |
| Neg | POS | Neg | Valid | hMPV detected. AdV and RV not detected. |
| Neg | Neg | POS | Valid | RV detected. AdV and hMPV not detected. |
| POS | POS | Neg | Valid | AdV and hMPV detected. RV not detected. |
| Neg | POS | POS | Valid | hMPV and RV detected. AdV not detected. |
| POS | Neg | POS | Valid | AdV and RV detected. hMPV not detected. |
| POS | POS | POS | Valid | AdV, hMPV and RV detected. Triple infections are rare. Retest to confirm result. |
| Invalid | Invalid | Invalid | Invalid | Invalid. There was an error in the generation of the result; retest sample. |

Note: POS result will be accompanied by cycle threshold (Ct) values.

Note: Detection of internal control is not required for samples that are positive for AdV, hMPV and/or RV.

Limitations

- A. Use of this assay is limited to personnel who are trained in the procedure. Failure to follow these instructions may result in erroneous results.
- B. Reliable results are dependent on adequate specimen collection, transport, storage, and processing.
- C. Avoid contamination by adhering to good laboratory practices and to the procedures specified in this package insert.
- D. Negative results do not preclude adenovirus, human metapneumovirus or rhinovirus infections and should not be used as the sole basis for treatment or other management decisions.
- E. There is a risk of hMPV or AdV false positive results in samples containing high Rhinovirus viral content. When there are dual-positive results with the Panther Fusion AdV/hMPV/RV assay where the RV positive result has a Ct value ≤ 26 with an RFU signal $\geq 28,000$ and the positive result for hMPV and/or AdV has a Ct value ≥ 39 , it is possible that the positive result for hMPV and/or AdV is a false positive result.
- F. This test does not differentiate Adenovirus subtypes (i.e. 1-58), human Metapneumovirus subtypes (i.e., A1, A2, B1, B2) or Rhinovirus species (i.e., Rhinovirus A, Rhinovirus B or Rhinovirus C); additional testing is required to differentiate any specific Adenovirus subtypes, human Metapneumovirus subtypes or specific Rhinovirus species in consultation with local public health departments.
- G. A positive result indicates the detection of nucleic acid from the relevant virus. Nucleic acid may persist even after the virus is no longer viable.

Panther Fusion System Assay Performance

Reproducibility

Panther Fusion AdV/hMPV/RV assay reproducibility was evaluated at three US sites using seven panel members. Testing was performed using one lot of assay reagents and six operators (two at each site). At each site, testing was performed for at least five days. Each run had three replicates of each panel member.

A negative panel member was created using a matrix of simulated nasal swab specimen in viral transport medium (VTM). Positive panel members were created by spiking 1-2X limit of detection (LoD, low-positive) or 2-3X LoD (moderate-positive) concentrations of the target analyte into a matrix of simulated nasal swab specimen, composed of cultured human cells suspended in VTM.

The agreement with expected results was 100% for all panel members containing AdV, hMPV, or RV, as shown in Table 2.

Table 2: Agreement of Panther Fusion AdV/hMPV/RV Assay Results with Expected Results

| Panel | | | Expected Results | | | Agreement with Expected Results | | | | | |
|--------------|-------------|--|------------------|------|----|---------------------------------|----------------|----------------|----------------|----------------|----------------|
| | | | | | | AdV | | hMPV | | RV | |
| Description | Composition | Concentration (TCID ₅₀ /mL) | AdV | hMPV | RV | N ¹ | (%) 95% CI | N ¹ | (%) 95% CI | N ¹ | (%) 95% CI |
| AdV Low Pos | 1-2X LoD | 1.00E+00 | + | - | - | 88/88 | 100 (95.8-100) | 88/88 | 100 (95.8-100) | 88/88 | 100 (95.8-100) |
| AdV Mod Pos | 2-3X LoD | 3.00E+00 | + | - | - | 89/89 | 100 (95.9-100) | 89/89 | 100 (95.9-100) | 89/89 | 100 (95.9-100) |
| hMPV Low Pos | 1-2X LoD | 1.00E+01 | - | + | - | 88/88 | 100 (95.8-100) | 88/88 | 100 (95.8-100) | 88/88 | 100 (95.8-100) |
| hMPV Mod Pos | 2-3X LoD | 3.00E+01 | - | + | - | 89/89 | 100 (95.9-100) | 89/89 | 100 (95.9-100) | 89/89 | 100 (95.9-100) |
| RV Low Pos | 1-2X LoD | 3.16E-01 | - | - | + | 89/89 | 100 (95.9-100) | 89/89 | 100 (95.9-100) | 89/89 | 100 (95.9-100) |
| RV Mod Pos | 2-3X LoD | 9.48E-01 | - | - | + | 87/87 | 100 (95.8-100) | 87/87 | 100 (95.8-100) | 87/87 | 100 (95.8-100) |
| Neg | N/A | N/A | - | - | - | 87/87 | 100 (95.8-100) | 87/87 | 100 (95.8-100) | 87/87 | 100 (95.8-100) |

CI= Score confidence interval, Mod= moderate, N/A= not applicable, Neg= negative, Pos= positive, TCID₅₀/mL= 50% tissue culture infective dose (measure of virus titer)

¹ A total of 13 samples had final invalid results and were not included in the calculation of overall agreement

The total AdV, hMPV, and RV signal variability measured as %CV ranged from 1.70% to 4.90% in low and moderate positive panel members. For the sources of variation except the 'within-run' factor, %CV values were ≤1.72% as shown in Table 3.

Table 3: Signal Variability of the Panther Fusion AdV/hMPV/RV Assay by Panel Member

| | | | Between Sites | | Between Operators | | Between Days | | Between Runs | | Within Runs | | Total | |
|-------------------|----|---------|---------------|--------|-------------------|--------|--------------|--------|--------------|--------|-------------|--------|-------|--------|
| Panel Description | N | Mean Ct | SD | CV (%) | SD | CV (%) | SD | CV (%) | SD | CV (%) | SD | CV (%) | SD | CV (%) |
| AdV Low Pos | 88 | 35.1 | 0.35 | 0.99 | 0.13 | 0.38 | 0.0 | 0.0 | 0.0 | 0.0 | 0.58 | 1.65 | 0.69 | 1.96 |
| AdV Mod Pos | 89 | 33.5 | <0.1 | 0.18 | 0.17 | 0.49 | 0.21 | 0.63 | <0.1 | <0.1 | 0.50 | 1.49 | 0.57 | 1.70 |
| hMPV Low Pos | 88 | 35.1 | 0.23 | 0.64 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 1.14 | 3.25 | 1.16 | 3.32 |
| hMPV Mod Pos | 89 | 33.1 | 0.0 | 0.0 | 0.24 | 0.71 | 0.57 | 1.72 | <0.1 | <0.1 | 1.50 | 4.53 | 1.62 | 4.90 |
| RV Low Pos | 89 | 33.7 | 0.14 | 0.43 | 0.24 | 0.72 | 0.22 | 0.66 | <0.1 | <0.1 | 0.83 | 2.45 | 0.90 | 2.67 |
| RV Mod Pos | 87 | 32.3 | 0.16 | 0.48 | <0.1 | 0.16 | 0.38 | 1.18 | <0.1 | 0.13 | 0.71 | 2.20 | 0.83 | 2.55 |

Ct= threshold cycle, CV= coefficient of variation, Mod= moderate, Pos= positive, SD= standard deviation

Note: If variability from some factors was numerically negative, SD and CV are shown as 0.0.

The signal variability, measured as %CV, was $\leq 1.94\%$ between sites, between operators, between days, or overall for the Panther Fusion AdV/hMPV/RV assay positive controls (see Table 4).

Table 4: Signal Variability of the Panther Fusion AdV/hMPV/RV Assay Controls

| | | | | Between Sites | | Between Operators | | Between Days | | Between Runs | | Within Runs | | Total | |
|---------|---------|----|---------|---------------|--------|-------------------|--------|--------------|--------|--------------|--------|-------------|--------|-------|--------|
| Control | Analyte | N | Mean Ct | SD | CV (%) | SD | CV (%) | SD | CV (%) | SD | CV (%) | SD | CV (%) | SD | CV (%) |
| Pos | AdV | 30 | 33.0 | 0.0 | 0.0 | 0.0 | 0.0 | <0.1 | 0.24 | 0.0 | 0.0 | 0.27 | 0.82 | 0.28 | 0.85 |
| | hMPV | 30 | 34.0 | <0.1 | 0.21 | <0.1 | 0.18 | 0.0 | 0.0 | 0.0 | 0.0 | 0.30 | 0.89 | 0.32 | 0.93 |
| | RV | 30 | 31.8 | 0.0 | 0.0 | 0.0 | 0.0 | 0.32 | 1.02 | 0.0 | 0.0 | 0.53 | 1.65 | 0.62 | 1.94 |

Ct= threshold cycle, CV= coefficient of variation, Pos= positive, SD= standard deviation

Note: If variability from some factors was numerically negative, SD and CV are shown as 0.0.

Clinical Performance

This study was performed to demonstrate clinical performance characteristics for the Panther Fusion AdV/hMPV/RV assay. A prospective multicenter study was conducted with leftover, remnant nasopharyngeal (NP) swab specimens from male and female individuals of all ages exhibiting signs and/or symptoms of a respiratory tract infection. Four participating US pediatric/adolescent, private and/or university hospitals obtained 2961 leftover, remnant NP swab specimens. The samples were tested with the Panther Fusion AdV/hMPV/RV assay, with reference viral culture followed by direct fluorescent antibody (DFA) identification (for AdV), with an FDA-cleared assay for hMPV, and with 2 reverse transcriptase PCR assays followed by bi-directional sequencing (PCR/sequencing, for RV). FDA-cleared or validated PCR-based assays were used for discordant resolution testing for AdV and hMPV; no discordant resolution testing was performed for RV. Performance characteristics were estimated relative to reference results for each sample. Sensitivity and specificity (for AdV and hMPV) and negative and positive percent agreement (for RV) were estimated with corresponding 2-sided 95% Score CIs. Analyses were performed separately for each target analyte (AdV, hMPV, RV).

Of the 2961 specimens, 31 specimens/samples were withdrawn (due to incomplete reference testing results, insufficient volumes for testing, expiration prior to testing, or mishandling), 2930 samples were processed in valid Panther Fusion AdV/hMPV/RV runs, 2875 (98.1%) had final valid results, and 55 (1.9%) had final invalid results. Of the 2875 samples with valid Panther Fusion results, 1358 samples were from females and 1517 samples were from males (see Table 5). Of the samples with valid Panther Fusion AdV/hMPV/RV results, 11 samples with invalid reference results for AdV (n=6) or RV (n=5) were excluded from the performance analyses, leaving 2869 samples evaluable for AdV, 2875 for hMPV, and 2870 for RV.

Table 5: Summary of Subject Demographics for Prospective Samples in the Panther Fusion AdV/hMPV/RV Assay Evaluation

| | | N (%) |
|-----------|----------------------|--------------|
| Total | | 2875 (100) |
| Sex | Female | 1358 (47.2) |
| | Male | 1517 (52.8) |
| Age Group | 0 to 28 days | 82 (2.9) |
| | 29 days to < 2 years | 757 (26.3) |
| | 2 to 5 years | 407 (14.2) |
| | 6 to 11 years | 259 (9.0) |
| | 12 to 17 years | 184 (6.4) |
| | 18 to 21 years | 73 (2.5) |
| | 22 to 64 years | 694 (24.1) |
| | ≥ 65 years | 419 (14.6) |

Of the 2875 evaluable samples tested using the Panther Fusion AdV/hMPV/RV assay, 5.6% (160/2869) were positive for AdV, 3.6% (103/2875) were positive for hMPV, and 21.0% (604/2870) were positive for RV. Table 6 shows the positivity for each analyte by age group.

Table 6: Panther Fusion AdV/hMPV/RV Assay Positivity by Analyte and Age Group

| Analyte | % Positivity (n/N) | | |
|----------------------|---------------------------|-----------------|------------------|
| | AdV | hMPV | RV |
| All | 5.6% (160/2869) | 3.6% (103/2875) | 21.0% (604/2870) |
| 0 to 28 days | 1.2% (1/82) | 1.2% (1/82) | 17.1% (14/82) |
| 29 days to < 2 years | 8.7% (66/757) | 5.8% (44/757) | 31.5% (238/756) |
| 2 to 5 years | 11.5% (47/407) | 6.9% (28/407) | 28.3% (115/406) |
| 6 to 11 years | 12.4% (32/258) | 2.3% (6/259) | 21.3% (55/258) |
| 12 to 17 years | 2.8% (5/181) | 0.5% (1/184) | 16.8% (31/184) |
| 18 to 21 years | 2.7% (2/73) | 1.4% (1/73) | 12.3% (9/73) |
| 22 to 64 years | 0.9% (6/692) | 2.2% (15/694) | 13.4% (93/692) |
| ≥ 65 years | 0.2% (1/419) | 1.7% (7/419) | 11.7% (49/419) |

Performance characteristics for detection of AdV, hMPV, and RV in prospective NP samples were calculated (see Table 7).

Table 7: Panther Fusion AdV/hMPV/RV Assay Performance Relative to Reference Testing

| Analyte | N | TP | FP | TN | FN | Prevalence ¹ (95% CI) ² | Sensitivity/PPA ³ (95% CI) ² | Specificity/NPA ³ (95% CI) ² |
|---------|------|-----|-----------------|------|-----------------|--|---|---|
| AdV | 2869 | 93 | 67 ⁴ | 2707 | 2 ⁴ | 3.3 (2.7-4.0) | 97.9 (92.6-99.4) | 97.6 (96.9-98.1) |
| hMPV | 2875 | 74 | 29 ⁵ | 2771 | 1 ⁵ | 2.6 (2.1-3.3) | 98.7 (92.8-99.8) | 99.0 (98.5-99.3) |
| RV | 2870 | 552 | 52 ⁶ | 2182 | 84 ⁶ | 22.2 (2.7-23.7) | 86.8 (83.9-89.2) | 97.7 (97.0-98.2) |

FN= false negative, FP= false positive, NPA= negative percent agreement, PPA= positive percent agreement, TP= true positive, TN= true negative

¹ Study prevalence reported, ² Score Confidence Interval, ³ PPA and NPA apply to RV.

⁴ 54/67 false positive results were confirmed positive and 2/2 false negative results were confirmed negative for AdV by an FDA-cleared assay.

⁵ 20/29 false positive results were confirmed positive and 0/1 false negative result was confirmed negative for hMPV by PCR.

⁶ No discordant resolution testing was performed for the 52 false positive and 84 false negative results for RV.

Analytical Sensitivity

The analytical sensitivity (limit of detection or LoD) of the Panther Fusion AdV/hMPV/RV assay for the NP swab specimen type was determined by testing pooled AdV/hMPV/RV negative clinical specimens spiked with the following virus cultures at various concentrations: Adenovirus (1, 3, 4, 9, 12, 40), hMPV (A1, A2, B1, B2) and RV (A-18 and B-26). At least twelve replicates were tested with each of the three reagent lots for a combined total of 36 replicates. Target specific LoD concentrations were verified by testing an additional 20 replicates with one reagent lot. Analytical sensitivity (LoD) is defined as the lowest concentration at which ≥95% of all replicates tested positive, as summarized in the table below.

Table 8: NP Swab Sensitivity

| Viral Strain | LoD Concentration |
|---------------------------|---|
| Adenovirus 1 (Species C) | 1x10 ⁰ TCID ₅₀ /mL |
| Adenovirus 3 (Species B) | 1x10 ⁰ TCID ₅₀ /mL |
| Adenovirus 4 (Species E) | 1x10 ⁻² TCID ₅₀ /mL |
| Adenovirus 9 (Species D) | 1x10 ^{-0.5} TCID ₅₀ /mL |
| Adenovirus 12 (Species A) | 1x10 ^{-0.5} TCID ₅₀ /mL |
| Adenovirus 40 (Species F) | 1x10 ^{-1.5} TCID ₅₀ /mL |
| hMPV A1-16 | 1x10 ² TCID ₅₀ /mL |
| hMPV A2-20 | 1x10 ¹ TCID ₅₀ /mL |
| hMPV B1-3 | 1x10 ^{0.5} TCID ₅₀ /mL |
| hMPV B2-8 | 1x10 ⁰ TCID ₅₀ /mL |
| Rhinovirus A-18 | 1x10 ^{-0.5} TCID ₅₀ /mL |
| Rhinovirus B-26 | 1x10 ⁰ TCID ₅₀ /mL |

Reactivity

The reactivity of the Panther Fusion AdV/hMPV/RV assay was evaluated against multiple strains of AdV, hMPV, and RV. Simulated reactivity evaluation was performed *in silico* for the types that are not available for testing. Reactivity was predicted for the AdV type 52-58 and RV type C.

Table 9: Reactivity Results

| Target | Description | Concentration | AdV | hMPV | RV |
|------------|-------------|--|-----|------|----|
| Adenovirus | AdV 1 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 2 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 3 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 4 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 5 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 6 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 7 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 8 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 9 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 10 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 11 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 12 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 13 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 14 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 15 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 16 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 17 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 19 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 20 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 21 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 22 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 23 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 24 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 25 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 26 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 27 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 28 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 29 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 30 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 31 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 32 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 33 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 34 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 35 | 1x10 ³ TCID ₅₀ /mL | + | - | - |

Table 9: Reactivity Results (continued)

| Target | Description | Concentration | AdV | hMPV | RV |
|--------------------------|--|--|-----|------|----|
| Adenovirus | AdV 36 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 37 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 38 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 39 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 40 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 41 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 42 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 43 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 44 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 45 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 46 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 47 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 48 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 49 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| | AdV 50 | 1x10 ³ TCID ₅₀ /mL | + | - | - |
| AdV 51 | 1x10 ³ TCID ₅₀ /mL | + | - | - | |
| human Metapneumovirus | hMPV A1-16 | 1x10 ³ TCID ₅₀ /mL | - | + | - |
| | hMPV A1-9 | 1x10 ³ TCID ₅₀ /mL | - | + | - |
| | hMPV A2-20 | 1x10 ³ TCID ₅₀ /mL | - | + | - |
| | hMPV A2-27 | 1x10 ³ TCID ₅₀ /mL | - | + | - |
| | hMPV B1-3 | 1x10 ³ TCID ₅₀ /mL | - | + | - |
| | hMPV B1-5 | 1x10 ³ TCID ₅₀ /mL | - | + | - |
| | hMPV B2-18 | 1x10 ³ TCID ₅₀ /mL | - | + | - |
| | hMPV B2-4 | 1x10 ³ TCID ₅₀ /mL | - | + | - |
| hMPV B2-8 | 1x10 ³ TCID ₅₀ /mL | - | + | - | |
| Rhinovirus* | RV A1 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A16 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A18 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A32 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A33 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A39 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A40 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A44 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A51 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A59 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A61 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A65 | 1x10 ² TCID ₅₀ /mL | - | - | + |

Table 9: Reactivity Results (continued)

| Target | Description | Concentration | AdV | hMPV | RV |
|-------------|-------------|--|-----|------|----|
| Rhinovirus* | RV A76 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A78 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A89 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV A100 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV B26 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV B52 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV B69 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV B70 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV B79 | 1x10 ² TCID ₅₀ /mL | - | - | + |
| | RV B86 | 1x10 ² TCID ₅₀ /mL | - | - | + |

* Simulated reactivity evaluation performed in-silico predicted reactivity with multiple Rhinovirus C strains.

Analytical Specificity

The analytical specificity of the Panther Fusion AdV/hMPV/RV assay was evaluated by testing a panel of 64 organisms, consisting of 30 viral, 32 bacterial, and 2 yeast strains representing common respiratory pathogens or flora commonly present in nasopharynx. Bacteria and yeast were tested at concentrations of 10⁵ to 10⁸ CFU/mL or IFU/mL, except where noted. Viruses were tested at concentrations of 10³ to 10⁷ TCID₅₀/mL.

Analytical specificity of the Panther Fusion AdV/hMPV/RV assay was 100% for AdV, hMPV, and RV.

Table 10: Specificity Results

| Organism | Concentration | AdV | hMPV | RV |
|---|--|-----|------|----|
| <i>Acinetobacter baumannii</i> 307-0294 | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Bordetella bronchiseptica</i> | 1x10 ⁷ CFU/ml | - | - | - |
| <i>Bordetella parapertussis</i> | 1x10 ⁷ CFU/ml | - | - | - |
| <i>Bordetella pertussis</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Burkholderia cepacia</i> Z066 | 1x10 ⁶ CFU/mL | - | - | - |
| <i>Candida albicans</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Candida glabrata</i> | 1x10 ⁶ CFU/mL | - | - | - |
| <i>Chlamydia pneumoniae</i> | 1x10 ⁵ CFU/mL | - | - | - |
| <i>Chlamydia trachomatis</i> | 1x10 ⁴ CFU/mL | - | - | - |
| CMV Strain AD 169 | 1x10 ⁴ TCID ₅₀ /mL | - | - | - |
| Coronavirus 229E | 1x10 ⁵ TCID ₅₀ /mL | - | - | - |
| Coronavirus OC43 | 1x10 ⁵ TCID ₅₀ /mL | - | - | - |
| <i>Corynebacterium diphtheria</i> | 1x10 ⁷ CFU ₅₀ /mL | - | - | - |
| Coxsackie B3 | 1x10 ⁶ TCID ₅₀ /mL | - | - | - |
| Coxsackie B4 | 1x10 ⁴ TCID ₅₀ /mL | - | - | - |

Table 10: Specificity Results (continued)

| Organism | Concentration | AdV | hMPV | RV |
|-------------------------------------|--|-----|------|----|
| Coxsackie B5/10/2006 | 1x10 ⁵ TCID ₅₀ /mL | - | - | - |
| Coxsackievirus A10 | 1x10 ⁴ TCID ₅₀ /mL | - | - | - |
| Coxsackievirus A21 | 1x10 ⁴ TCID ₅₀ /mL | - | - | - |
| <i>E. coli</i> | 1x10 ⁷ CFU/mL | - | - | - |
| EBV | 1x10 ⁶ TCID ₅₀ /mL | - | - | - |
| Echovirus 11 | 1x10 ⁶ TCID ₅₀ /mL | - | - | - |
| Echovirus 2 | 1x10 ⁶ TCID ₅₀ /mL | - | - | - |
| Echovirus 3 | 1x10 ⁴ TCID ₅₀ /mL | - | - | - |
| Echovirus 6 | 1x10 ⁶ TCID ₅₀ /mL | - | - | - |
| Enterovirus 68 | 1x10 ⁵ TCID ₅₀ /mL | - | - | - |
| Enterovirus 70 | 1x10 ⁴ TCID ₅₀ /mL | - | - | - |
| Haemophilus Influenzae | 1x10 ⁷ TCID ₅₀ /mL | - | - | - |
| HPIV-1 | 1x10 ⁴ TCID ₅₀ /mL | - | - | - |
| HPIV-2 | 1x10 ⁵ TCID ₅₀ /mL | - | - | - |
| HPIV-3 | 1x10 ⁵ TCID ₅₀ /mL | - | - | - |
| HPIV-4a | 1x10 ⁴ TCID ₅₀ /mL | - | - | - |
| HSV-1 Macintyre Strain | 1x10 ⁵ TCID ₅₀ /mL | - | - | - |
| HSV-2 Type 2G Strain | 1x10 ⁵ TCID ₅₀ /mL | - | - | - |
| Influenza A (H1N1) | 1x10 ³ TCID ₅₀ /mL | - | - | - |
| Influenza A (H3N2) | 1x10 ³ TCID ₅₀ /mL | - | - | - |
| Influenza B | 1x10 ³ TCID ₅₀ /mL | - | - | - |
| <i>Klebsiella pneumonia</i> | 1x10 ⁷ CFU/mL | - | - | - |
| Lactobacillus acidophilus Z048 | 1x10 ⁶ CFU/mL | - | - | - |
| <i>Lactobacillus plantarum</i> | 1x10 ⁶ CFU/mL | - | - | - |
| <i>Legionella pneumophila</i> | 1x10 ⁷ CFU/mL | - | - | - |
| Measles/7/2000 | 1x10 ⁴ TCID ₅₀ /mL | - | - | - |
| <i>Moraxella catarrhalis</i> | 1x10 ⁷ CFU/mL | - | - | - |
| Mumps virus | 1x10 ⁵ CFU/mL | - | - | - |
| <i>Mycobacterium intracellulare</i> | 5x10 ¹⁰ rRNA copies/mL | - | - | - |
| <i>Mycobacterium tuberculosis</i> | 5x10 ⁹ rRNA copies/mL | - | - | - |
| <i>Mycoplasma pneumoniae</i> | 1x10 ⁶ CFU/mL | - | - | - |
| <i>Neisseria gonorrhoea</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Neisseria meningitides</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Neisseria mucosa</i> | 1x10 ⁷ CFU/mL | - | - | - |
| Polio virus 1 | 1x10 ⁶ TCID ₅₀ /mL | - | - | - |

Table 10: Specificity Results (continued)

| Organism | Concentration | AdV | hMPV | RV |
|---|--|-----|------|----|
| <i>Proteus mirabilis</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Proteus vulgaris</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Pseudomonas aeruginosa</i> | 1x10 ⁷ CFU/mL | - | - | - |
| RSV A | 1x10 ⁵ TCID ₅₀ /mL | - | - | - |
| RSV B | 1x10 ⁵ TCID ₅₀ /mL | - | - | - |
| <i>Serratia marcescens</i> Z053 | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Staphylococcus aureus</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Staphylococcus epidermidis</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Streptococcus agalactiae</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Streptococcus pneumoniae</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Streptococcus pyogenes</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Streptococcus salivarius</i> | 1x10 ⁷ CFU/mL | - | - | - |
| <i>Tatlockia micdadei</i> (<i>Legionella micdadei</i>) | 1x10 ⁶ CFU/mL | - | - | - |
| Varicella Zoster Virus | 1x10 ⁴ TCID ₅₀ /mL | - | - | - |

Competitive Interference

Competitive Interference of the Panther Fusion AdV/hMPV/RV assay was evaluated using a simulated clinical matrix with pairs of target viruses at two different concentrations. One of the concentrations was near the Limit of Detection (3X LoD) while the other concentration was high (1000X LoD). The presence of two viruses at varying concentrations in a single sample had no effect on the analytical sensitivity (100% detection for both targets) at the concentration noted in the table below.

Table 11: Competitive Interference

| Condition | Target 1 | | Target 2 | | AdV Result | hMPV Result | RV Result |
|-----------|-------------|---------------|-------------|---------------|------------|-------------|-----------|
| | Description | Concentration | Description | Concentration | | | |
| 1 | AdV | 3X LoD | hMPV | 1000X LoD | + | + | - |
| 2 | AdV | 3X LoD | RV | 1000X LoD | + | - | + |
| 3 | hMPV | 3X LoD | AdV | 1000X LoD | + | + | - |
| 4 | hMPV | 3X LoD | RV | 1000X LoD | - | + | + |
| 5 | RV | 3X LoD | AdV | 1000X LoD | + | - | + |
| 6 | RV | 3X LoD | hMPV | 1000X LoD | - | + | + |

Interference

Mucin, whole blood and other potentially interfering substances (medications and over-the-counter or OTC products) that may be present in the samples were evaluated in the Panther Fusion AdV/hMPV/RV assay. Clinically relevant amount of the potentially interfering substances were added to simulated clinical matrix and tested unspiked or spiked with cultured AdV, hMPV

and RV at their respective 3X LoD concentrations. The substances consisted of nasal sprays (liquid and powder), ingestible pills, lozenges, injectable and endogenous substances, as shown in Table 12.

All of the substances tested were found to have no impact on the performance of the Panther Fusion AdV/hMPV/RV assay, at the concentrations tested.

Table 12: Potentially Interfering Substances

| Type | Substance Name | Active Ingredient(s) | Concentration |
|----------------------------|------------------------------|--|---------------|
| Endogenous | Mucin | Purified mucin protein | 60 µg/mL |
| | Human blood | Blood | 2% v/v |
| Nasal sprays or drops | Neo-Synephrine® | Phenylephrine | 15% v/v |
| | Anefrin | Oxymetazoline | 15% v/v |
| | Saline | Sodium chloride | 15% v/v |
| | Ventolin® HFA | Albuterol | 15% v/v |
| Nasal corticosteroids | QVAR®, Beconase AQ | Beclomethasone | 5% v/v |
| | Dexacort | Dexamethasone | 5% v/v |
| | AEROSPAN® | Flunisolide | 5% v/v |
| | Nasacort | Triamcinolone | 5% v/v |
| | Rhinocort | Budesonide | 5% v/v |
| | Nasonex | Mometasone | 5% v/v |
| | Flonase | Fluticasone | 5% v/v |
| Nasal gel | Zicam® (Allergy Relief) | Luffa operculata, Galphimia, Glauca, Histaminum hydrochloricum, Sulfur | 5% v/v |
| Throat lozenges | Chloraseptic Throat Lozenges | Benzocaine Menthol | 0.63 mg/mL |
| Anti-viral drugs | Relenza® | Zanamivir | 3.3 mg/mL |
| | TamiFlu | Oseltamivir | 25 mg/mL |
| | Rebitol | Ribavirin | 20 mg/mL |
| Antibiotic, nasal ointment | Bactroban cream | Mupirocin | 10 mg/mL |
| Antibiotic, systemic | Tobramycin | Tobramycin | 4.0 µg/mL |

Carry-Over/Contamination

The carry-over/cross-contamination study was performed with negative samples alternately placed between high positive samples and tested. High positive samples were prepared by spiking (over 10,000X LoD). Nine separate runs with negative samples and positive samples placed in a checkerboard pattern were tested over three different instruments for a combined total of 449 positive and 450 negative samples. The carry-over rate was 0.2%.

Assay Precision

Panther Fusion AdV/hMPV/RV assay precision was evaluated with a 7-member panel. The panel was tested by three operators on two separate runs per day, using three reagent lots on three Panther Fusion systems over 45 days.

The panel members are described in Table 13, along with a summary of the agreement with expected results for each targets. Table 14 presents the mean and variability analysis between instruments, between reagent lots, between operators, between days, between runs and within runs, and overall (total) for Ct.

Table 13: Panel Description and % Agreement

| Target | Panel Member | % Positive | % Total Agreement (95% CI) |
|--------|-------------------|---------------------|----------------------------|
| AdV | AdV 3X LoD | 100.0% (162/162) | 100.0% (97.7 - 100%) |
| | AdV 1X LoD | 100.0% (162/162) | 100.0% (97.7 - 100%) |
| | AdV 0.01X LoD | 10.6% (17/161) | 89.4% (83.7 - 93.3%) |
| | Negative | 0.6% (1/162) | 99.4% (96.6 - 99.9%) |
| hMPV | hMPV 3X LoD | 100.0% (160/160) | 100.0% (97.7 - 100%) |
| | hMPV 1X LoD | 100.0% (161/161) | 100.0% (97.7 - 100%) |
| | hMPV 0.01X LoD | 17.9% (29/162) | 82.1% (75.5 - 87.2%) |
| | Negative | 0.0% (0/162) | 100.0% (97.7% - 100.0%) |
| RV | RV 3X LoD | 100.0% (161/161) | 100.0% (97.7 - 100%) |
| | RV 1X LoD | 100.0% (162/162) | 100.0% (97.7 - 100%) |
| | RV 0.01X LoD | 1.9% (3/160) | 98.1% (94.6 - 99.4%) |
| | Negative | 0.6% (1/162) | 99.4% (96.6 - 99.9%) |

Table 14: Signal Variability

| Target | Panel Member | Mean Ct | Between Instrument | | Between Reagent Lots | | Between Operator | | Between Days | | Between Runs | | Within Runs | | Total | |
|--------|----------------|---------|--------------------|--------|----------------------|--------|------------------|--------|--------------|--------|--------------|--------|-------------|--------|-------|--------|
| | | | SD | CV (%) | SD | CV (%) | SD | CV (%) | SD | CV (%) | SD | CV (%) | SD | CV (%) | SD | CV (%) |
| AdV | AdV 3X LoD | 33.5 | 0.1 | 0.4 | 0.0 | 0.1 | 0.0 | 0.0 | 0.1 | 0.3 | 0.2 | 0.7 | 0.4 | 1.2 | 0.5 | 1.5 |
| | AdV 1X LoD | 35.2 | 0.2 | 0.6 | 0.0 | 0.0 | 0.0 | 0.2 | 0.1 | 0.3 | 0.3 | 0.8 | 0.5 | 1.5 | 0.6 | 1.9 |
| | AdV 0.01X LoD | 40.4 | 0.3 | 0.8 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.9 | 2.4 | 0.7 | 1.9 | 1.3 | 3.2 |
| hMPV | hMPV 3X LoD | 33.5 | 0.0 | 0.1 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.2 | 0.8 | 0.8 | 2.4 | 0.8 | 2.5 |
| | hMPV 1X LoD | 35.1 | 0.0 | 0.1 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.1 | 0.7 | 2.0 | 0.7 | 2.0 |
| | hMPV 0.01X LoD | 40.3 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.3 | 0.7 | 0.5 | 1.4 | 1.2 | 3.1 | 1.4 | 3.5 |
| RV | RV 3X LoD | 32.5 | 0.1 | 0.5 | 0.1 | 0.3 | 0.0 | 0.1 | 0.0 | 0.0 | 0.3 | 1.0 | 0.6 | 2.0 | 0.7 | 2.4 |
| | RV 1X LoD | 33.8 | 0.1 | 0.5 | 0.1 | 0.5 | 0.0 | 0.0 | 0.1 | 0.4 | 0.0 | 0.0 | 0.8 | 2.6 | 0.9 | 2.8 |
| | RV 0.01X LoD | 40.6 | 1.9 | 4.7 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.6 | 1.6 | 2.0 | 5.0 |
| IC | Negative | 30.7 | 0.0 | 0.2 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.0 | 0.1 | 0.6 | 0.5 | 1.7 | 0.5 | 1.8 |

Bibliography

1. Centers for Disease Control and Prevention. National Respiratory and Enteric Virus Surveillance System. Centers for Disease Control and Prevention Web site. <http://www.cdc.gov/surveillance/nrevss/>. Accessed October, 2015.
2. Kahn, J.S. 2006. Epidemiology of human metapneumovirus. *Clin. Microbiol. Rev.* 19:546-557.
3. <http://www.cdc.gov/adenovirus/hcp/clinical-overview.html>. Accessed June 2016.
4. Martin, Malcolm A.; Knipe, David M.; Fields, Bernard N.; Howley, Peter M.; Griffin, Diane; Lamb, Robert (2007). *Fields' virology*. Philadelphia: Wolters Kluwer Health/Lippincott Williams & Wilkins. p. 2395.
5. <http://www.cdc.gov/adenovirus/outbreaks.html>. Accessed June 2016.
6. Kahn, J.S., Epidemiology of human metapneumovirus. *Clin Microbiol Rev*, 2006. 19(3): p. 546-57.
7. <http://www.cdc.gov/surveillance/nrevss/hmpv/clinical.html>. Accessed June 2016.
8. Park, J. Y., Yun, K. W., Lim, J. W., Lee, M. K., Lim, I. S., and Choi, E. S. (2016) Clinical and genetic features of human metapneumovirus infection in children. *Pediatrics International*, 58: 22–26. doi: 10.1111/ped.12782.
9. Anzueto, A. and M.S. Niederman. 2003. Diagnosis and treatment of rhinovirus respiratory infections. *Chest* 123:1664-1672.
10. Clinical & Laboratory Standards Institute. Document M29 Protection of Laboratory Workers from Occupationally Acquired Infections. CLSI Web site <https://clsi.org/standards/products/microbiology/documents/m29/>. Accessed September 2017.



Hologic, Inc.
10210 Genetic Center Drive
San Diego, CA 92121 USA

Customer Support: +1 800 442 9892
customersupport@hologic.com

Technical Support: +1 888 484 4747
molecularsupport@hologic.com

For more contact information visit www.hologic.com.

Hologic, Aptima, and Panther Fusion are trademarks and/or registered trademarks of Hologic, Inc. and/or its subsidiaries in the United States and/or other countries.

All other trademarks that may appear in this package insert are the property of their respective owners.

This product may be covered by one or more U.S. patents identified at www.hologic.com/patents.

©2019 Hologic, Inc. All rights reserved.

AW-16834 Rev. 003
2019-10