

Date:01/27/2021Author:Software EngineeringSystem:ApexSubsystem:Microsoft Windows OS 10Subject:Microsoft Windows 10 OS Patch release January 2021

Purpose:

To release the list of validated Microsoft critical patches on the Apex systems.

Note 1: If you are using the following versions of APEX, if you are required to update to the latest Hologic validated Windows updates, you must ensure your system(s) have been patched with the "Sentinel Run-Time 7.80.0.0" which can be located in the Windows Control Panel "Program and Features". If needed you can install the patch by ordering "APEX Service Dongle Patch Release" (ASY-12610). Please contact Hologic Support if you have any questions.

System Type	Software Versions	
Horizon	APEX v5.6.0.7	
Workstation	APEX v5.6.0.7	

List of patches/KB's validated for January 2021:

KB4598243 KB4535680

List of all patches/KB's validated on the Apex System:

KB4598243	KB4565511	KB4537764
KB4593226	KB4562561	KB4537759
KB4586830	KB4561616	KB4535680
KB4580346	KB4561600	KB4535288
KB4580325	KB4556813	KB4534271
KB4577015	KB4550994	KB4530689
KB4576750	KB4550929	KB4520724
KB4571694	KB4540723	KB4516115
KB4565912	KB4540670	KB4516044

KB4512574	KB4498947	KB4018073
KB4505422	KB4494175	KB2565063
KB4503537	KB4346087	KB2467173
KB4503308	KB4132216	
KB4503267	KB4022619	