

Date: 08/06/2020

Author: Software Engineering

System: Apex

Subsystem: Microsoft Windows OS 10

Subject: Microsoft Windows 10 OS Patch release July 2020

Purpose:

To release the list of validated Microsoft critical patches on the Apex systems.

Note 1: If you are using the following versions of APEX, if you are required to update to the latest Hologic validated Windows updates, you must ensure your system(s) have been patched with the "Sentinel Run-Time 7.80.0.0" which can be located in the Windows Control Panel "Program and Features". If needed you can install the patch by ordering "APEX Service Dongle Patch Release" (ASY-12610). Please contact Hologic Support if you have any questions.

System Type	Software Versions
Horizon	APEX v5.6.0.7
Workstation	APEX v5.6.0.7

List of patches/KB's validated for July 2020:

KB4565912 KB4565511 KB4535288

List of all patches/KB's validated on the Apex System:

KB4565912 KB4540723 KB4132216

KB4565511 KB4535288 KB4022619

KB4561600 KB4494175