



Date: 10/15/2020

Author: Software Engineering

System: Apex

Subsystem: Microsoft Windows OS 10

Subject: Microsoft Windows 10 OS Patch release September 2020

Purpose:

To release the list of validated Microsoft critical patches on the Apex systems.

Note 1: If you are using the following versions of APEX, if you are required to update to the latest Hologic validated Windows updates, you must ensure your system(s) have been patched with the "Sentinel Run-Time 7.80.0.0" which can be located in the Windows Control Panel "Program and Features". If needed you can install the patch by ordering "APEX Service Dongle Patch Release" (ASY-12610). Please contact Hologic Support if you have any questions.

System Type	Software Versions
Horizon	APEX v5.6.0.7
Workstation	APEX v5.6.0.7

List of patches/KB's validated for September 2020:

KB4577015 KB4576750

List of all patches/KB's validated on the Apex System:

KB4577015	KB4550994	KB4520724
KB4576750	KB4550929	KB4516115
KB4571694	KB4540723	KB4516044
KB4565912	KB4540670	KB4512574
KB4565511	KB4537764	KB4505422
KB4562561	KB4537759	KB4503537
KB4561616	KB4535288	KB4503308
KB4561600	KB4534271	KB4503267
KB4556813	KB4530689	KB4498947

KB4494175

KB4022619

KB2467173

KB4346087

KB4018073

KB4132216

KB2565063