



Genfind[®] DNA Extraction Kit

REF 95-449

Intended Use: Kit for DNA Extraction



TABLE OF CONTENTS

INTENDED USE
REAGENTS PROVIDED AND STORAGE REQUIREMENTS
WARNINGS AND PRECAUTIONS
LIMITATIONS
MATERIALS REQUIRED, BUT NOT PROVIDED
PREPARATION OF REAGENTS
INSTRUCTIONS FOR USE
PROCEDURAL NOTES AND PRECAUTIONS

INTENDED USE

The Genfind[®] DNA Extraction Kit is intended for use in the extraction of DNA from cervical specimens collected in ThinPrep[®] Pap Test PreservCyt[®] Solution for testing by the Cervista[®] HPV HR and Cervista[®] HPV 16/18 tests.

REAGENTS PROVIDED AND STORAGE REQUIREMENTS

Note: For information on any hazard and precautionary statements that may be associated with reagents, refer to the Safety Data Sheet Library at www.hologic.com/sds.

Table 1: Genfind DNA Extraction Kit (REF 95-449) Contents and Storage Requirements

Reagent	Label Abbreviation	Component Description	Storage Requirement
Genfind Proteinase K	PK	Lyophilized Enzyme (1 mL vials) Ultrapure	-30° to -15°C Store Frozen
Genfind Lysis Buffer	LB	Cell Lysis Solution 0.45 µm filtered	15° to 30°C Store at room temperature
Genfind Binding Buffer	BB	Magnetic Bead Solution 0.45 µm filtered	2° to 8°C Refrigerate – Do Not Freeze
Genfind Wash Buffer	WB	DNA Wash Buffer (Label marked with blue stripes) 0.45 µm filtered	15° to 30°C Store at room temperature

WARNINGS AND PRECAUTIONS

1. Universal safety precautions should be used when handling any human tissues or fluids. Specimens should be disposed of according to local requirements.
2. Product components (product residuals, packaging) can be considered as laboratory waste. Dispose of unused reagents and waste in accordance with applicable federal, state, and local regulations.
3. Sodium azide forms explosive compounds with heavy metals. This product contains concentrations of azide <0.1% (w/w) which with repeated contact with lead and copper commonly found in plumbing drains may result in the build up of shock-sensitive compounds.

LIMITATIONS

1. Genfind® DNA Extraction Kit is not intended for use with any sample type other than cervical samples collected in ThinPrep Pap Test PreservCyt Solution.
2. Genfind DNA Extraction Kit is not intended for use with any assay other than the Cervista HPV HR and Cervista HPV 16/18 tests.

MATERIALS REQUIRED, BUT NOT PROVIDED

Consumable Supplies

- Pipette tips, filter barrier
- 96-well plates
- Foil plate sealers
- ABgene® 96-well 2.2 mL plates
- Appropriately-sized nuclease-free disposable tubes and screw caps

Reagents

- 2M Tris, pH 7.5
- Nuclease-free water
- 70% Ethanol (Molecular Biology grade)

Equipment

- Cervista® HTA System for automation users
- Pipettes
- Vortex
- Plate centrifuge and rotors
- SPRI® Plate 96R Super Magnet Plate
- Thermomixer R (Eppendorf)
- MTP Block (Eppendorf) and 96-well adapter plate

PREPARATION OF REAGENTS

Equilibrate all reagents to room temperature prior to use. Prepared reaction mixtures should not be stored for later use.

1. Prepare a 10 mM Tris, pH 7.5, solution from a 2M Tris stock solution. For processing a 96-well plate of samples, a recommended preparation is shown in Table 2.

Table 2: Preparation of 10 mM Tris, pH 7.5.

Component	Volume
2M Tris, pH 7.5	100 μ L
Nuclease-free water	19.9 mL
Total Solution Volume	20 mL

2. Combine the Lysis Buffer and Proteinase K (96 μ g/ μ L) in an appropriate-sized conical tube according to Table 3. Mix by pipetting up and down.

Table 3: Preparation of Lysis Buffer.

Component	Volume/ Sample	Number of Samples (x)	Total Volume
Lysis Buffer	400 μ L	x	(400 μ L)(x)(1.2)
Proteinase K	9 μ L	x	(9 μ L)(x)(1.2)
LB/PK solution	409 μ L	x	(409 μ L)(x)(1.2)

INSTRUCTIONS FOR USE FOR CERVISTA HTA SYSTEM SYSTEM

Refer to the Cervista HTA Operator's Manual (P/N MAN-01522-001) for the instructions for use for the Cervista HTA System.

INSTRUCTIONS FOR USE FOR CERVISTA MANUAL SYSTEM

1. Mix the PreservCyt cervical specimen well by vortexing or shaking vigorously. Transfer 2.0 mL of each specimen to a well of a 96-well 2.2 mL plate.
2. Centrifuge the 96-well 2.2 mL plate at 1107 RCF for 10-15 minutes.
3. Place the 2.2 mL plate on the SPRI Plate 96R Super Magnet Plate. Remove the supernatant using a multi-channel pipette or a 96-well aspirator and pump (pressure of aspirator should be approximately 100 mm Hg vac). Remove approximately 1.9 mL of the supernatant leaving

- 50-100 μL of residual volume. Take care to only remove supernatant and not cellular material. **NOTE: IF USING AN ASPIRATOR RINSE WITH FRESH DISTILLED WATER FOLLOWING STEPS 3, 8, 11, 12, and 14).**
4. Add 400 μL of the Lysis Buffer/Proteinase K mixture to each well containing sample of the 96-well plate. **NOTE: USE NEW TIPS FOR EACH SAMPLE WELL IN ALL LIQUID TRANSFER STEPS.**
 5. Incubate the plate on a thermomixer for 15 minutes at 37°C +/-2°C and 1000 rpm. **NOTE: AFTER THIS STEP, TURN THE THERMOMIXER THERMOSTAT OFF. THE THERMOMIXER THERMOSTAT SHOULD REMAIN OFF FOR ALL SUBSEQUENT STEPS.**
 6. **IMPORTANT:** Mix the Binding Buffer thoroughly by inverting the bottle many times, making sure the beads are fully resuspended. After mixing, add 200 μL to each well containing sample of the 96-well plate.
 7. Place the plate on a thermomixer and mix at 1000 rpm for 2-3 minutes.
 8. Place the SPRI Plate 96R Super Magnet Plate on the spacer and place the 2.2 mL plate on the magnet for 4-6 minutes or until beads form a distinct ring and solution is clear. Aspirate the entire supernatant taking care not to disturb the beads. **NOTE: USE OF A SPACER IS NECESSARY FOR ALL SUBSEQUENT ASPIRATION STEPS IF USING THE 96-WELL ASPIRATOR AND PUMP.**
 9. Remove the plate from the magnet and spacer and add 400 μL of Wash Buffer to the plate wells containing beads.
 10. Place the plate on a thermomixer and mix at 1000 rpm for 4-6 minutes.
 11. Place the SPRI Plate 96R Super Magnet Plate on the spacer and place the 2.2 mL plate on the magnet for 4-6 minutes or until beads form a distinct ring and solution is clear. Aspirate entire supernatant taking care to not disturb the beads. **NOTE: THE PLATE SHOULD REMAIN ON THE MAGNET AND SPACER DURING STEPS 12-14.**
 12. Add 400 μL of 70% ethanol to the wells containing beads and incubate for 30-60 seconds. The beads should form a distinct ring. Aspirate entire supernatant.
 13. Repeat the 70% ethanol wash by adding 400 μL 70% of ethanol to the wells containing beads and incubate for 30-60 seconds. The beads should form a distinct ring. Aspirate entire supernatant.
 14. Allow the beads to air dry for 3-4 minutes. **NOTE: IT IS IMPORTANT TO REMOVE ALL RESIDUAL ETHANOL BEFORE PROCEEDING TO THE NEXT STEP.**
 15. Remove the plate from the magnet and add 120 μL of 10 mM Tris-HCl, pH 7.5 to each well containing beads.

16. Place the plate on a thermomixer and alternate mixing at:
 - i. 1000 rpm for 2-3 minutes
 - ii. Let stand for 2-3 minutes.
 - iii. 1000 rpm for 2-3 minutes.
17. Place the plate on a magnet for 10 minutes or until beads form a distinct ring and solution is clear.
18. While the plate is still on the magnet, transfer 110 μL of the DNA solution to a clean 96-well PCR plate using a multi-channel pipette.
19. If the beads are present visually in the DNA solution, place the 96-well PCR plate on the magnet and allow any particles to settle. While the plate is still on a magnet, transfer 100 μL of DNA to a clean 96-well PCR plate. Seal the plate with a foil plate sealer.
20. DNA can be stored at 4–8°C for up to four weeks. For storage longer than four weeks, store the sample DNA in a –20° or –80°C non-frost-free freezer.

PROCEDURAL NOTES AND PRECAUTIONS

For *in vitro* diagnostic use.

1. Multiple storage conditions exist; see Table 1.
2. Follow good laboratory practices. Wear protective disposable gloves, laboratory coats, and eye protection when handling specimens and kit reagents. Wash hands thoroughly after handling specimens and reagents.
3. Do not pool reagents from different lots or from different vials/bottles of the same lot.
4. Do not use reagents after their expiration date.
5. Prior to use, the Proteinase K lyophilized enzyme should be dissolved in nuclease-free water. A volume of 1 mL of water should be added to each vial as needed. When resuspended with water, the 1 mL vial of Proteinase K should be divided into aliquots and again frozen at -30° to -15°C in a non-frost-free freezer. Thaw only as much Proteinase K as needed for each extraction. Repeated freezing and thawing of the enzyme can cause a loss of function.
6. If a white precipitate has formed in the Wash Buffer, prior to use, gently shake or stir at room temperature until the solids dissolve. Do not heat to recombine.

Contact Information:

Manufacturer:

Hologic, Inc.
10210 Genetic Center Drive
San Diego, CA 92121 USA
Customer Support: +1 844 Hologic (+1 844 465 6442)
customersupport@hologic.com

Technical Support: +1 888 484 4747
molecularsupport@hologic.com

For more contact information visit www.hologic.com.

NOTICE TO RECIPIENT ABOUT LIMITED LICENSE

The Genfind® DNA Extraction Kit utilizes SPRI® paramagnetic bead technology and additional components, covered under U.S. Patent Nos. 5,705,628; 5,898,071; 6,534,262 and any corresponding international equivalents.

LIMITED PRODUCT WARRANTY

WARRANTIES. Equipment, Supplies, and Software are warranted to the original Customer to perform substantially in accordance with published Product Specifications for one (1) year starting from the date of Installation (if applicable) or from the date of Delivery, whichever occurs first. After-sale options and accessories are warranted for six (6) months, and x-ray tubes are warranted on a straight-line prorated basis as stated in the applicable Product Specification ("Warranty Period"). Replacement parts are warranted for the remainder of the Warranty Period or ninety (90) days from Delivery, whichever is longer. Consumable Supplies are warranted to conform to published specifications for a period ending on the expiration date shown on their respective packages. Services are warranted to be supplied in a workman-like manner. Hologic does not warrant that use of Products will be uninterrupted or error-free, or that Products will operate with non-Hologic authorized third-party products. HOLOGIC'S ENTIRE WARRANTY RESPONSIBILITY IS EXPRESSLY LIMITED TO REPAIR OR REPLACEMENT (AT HOLOGIC'S OPTION AND IN THE FORM ORIGINALLY SHIPPED) OF PRODUCT OR CORRECTION OF SERVICE SUBJECT TO ANY CLAIM, OR, AT HOLOGIC'S ELECTION, REPAYMENT OF, OR CREDITING CUSTOMER WITH, AN AMOUNT EQUAL TO THE HOLOGIC PRICE, FEE OR CHARGE THEREFORE. THE FOREGOING WARRANTIES ARE IN LIEU OF AND EXCLUDE ALL OTHER WARRANTIES NOT EXPRESSLY SET FORTH HEREIN, WHETHER EXPRESS OR IMPLIED BY OPERATION OF LAW OR OTHERWISE, INCLUDING BUT NOT LIMITED TO ANY IMPLIED WARRANTIES OF MERCHANTABILITY OR FITNESS FOR A PARTICULAR

PURPOSE. SUCH LIMITED WARRANTY IS GIVEN SOLELY TO THE ORIGINAL CUSTOMER AND IS NOT GIVEN TO, NOR MAY IT BE RELIED UPON BY, ANY THIRD PARTY INCLUDING, WITHOUT LIMITATION, CUSTOMERS OF CUSTOMER. THIS WARRANTY IS VOID UPON TRANSFER OF PRODUCT BY CUSTOMER TO ANY ENTITY WHO HAS LESS THAN FIFTY (50) PERCENT OWNERSHIP IN THE PRODUCT. SOME STATES DO NOT ALLOW THE EXCLUSION OF IMPLIED WARRANTIES SO THE ABOVE EXCLUSIONS MAY NOT APPLY TO YOU. YOU MAY ALSO HAVE OTHER RIGHTS, WHICH VARY, FROM STATE TO STATE. These warranties do not apply to any item that is: (a) repaired, moved or altered other than by Hologic authorized service personnel; (b) subjected to physical (including thermal or electrical) abuse, stress, or misuse; (c) stored, maintained or operated in any manner inconsistent with applicable Hologic specifications or instructions; or (d) designated as supplied subject to a non-Hologic warranty or on a pre-release or "as-is" basis.

WARRANTY CLAIMS AND REMEDIES. In the event of any warranty claim, Hologic will replace with new or repaired items any Equipment part, component, or consumable supply that is in breach of warranty, and will use reasonable efforts to promptly fix or provide a workaround for any Software defect or bug which prevents operation in substantial conformity with functional specifications. Alternatively, Hologic may elect to repay or credit to Customer an amount equal to the purchase price of the defective Equipment, component, Software, consumable supply, or Service. Items replaced shall become Hologic property. All claims shall be initiated by contacting Hologic within the applicable warranty period and thirty (30) days after discovery of the breach or non-conformity. Hologic must be given reasonable access and an opportunity to inspect all associated materials. If Hologic and Customer are unable to settle any claim and Customer has not notified Hologic within one (1) year after the claim arises, Customer shall be barred from instituting any legal action thereafter. These remedies shall comprise Hologic's entire liability and Customer's exclusive remedy for breach of warranty and are in lieu of any other remedies at law or equity.

LIMIT OF LIABILITY. HOLOGIC SHALL NOT BE LIABLE FOR ANY SPECIAL, INCIDENTAL, PUNITIVE, EXEMPLARY OR CONSEQUENTIAL LOSSES, DAMAGES, OR EXPENSES (INCLUDING BUT NOT LIMITED TO LOSS OF PROFITS, DATA, OR USE), DIRECTLY OR INDIRECTLY ARISING FROM THE SALE, HANDLING, SERVICE OR USE OF PRODUCT ORDERED OR FURNISHED, OR FROM ANY CAUSE RELATING THERETO UNLESS EXPRESSLY AGREED TO BY THE PARTIES IN WRITING. EXCEPT FOR PERSONAL INJURY OR DEATH TO THE EXTENT RESULTING FROM HOLOGIC'S NEGLIGENT OR INTENTIONALLY WRONGFUL ACTS OR OMISSIONS, IN NO EVENT SHALL HOLOGIC BE

LIABLE UNDER ANY LEGAL THEORY OR FOR ANY CAUSE WHATSOEVER, WHETHER BASED UPON WARRANTY, CONTRACT, TORT, NEGLIGENCE, OR OTHER THEORY, EVEN IF ADVISED OF THE POSSIBILITY THEREOF, FOR ANY AMOUNT IN EXCESS OF THE PRICE, FEE OR CHARGE THEREFORE RECEIVED BY HOLOGIC.

Hologic, Cervista, PreservCyt, and ThinPrep are trademarks and/or registered trademarks of Hologic, Inc. and/or its subsidiaries in the United States and/or other countries.

Genfind and SPRI are registered trademarks of Beckman Coulter Genomics.

All other trademarks that may appear in this package insert are the property of their respective owners.

©2010-2015 Hologic, Inc. All rights reserved.

P/N 15-3110, Revision 103